

Service Repair Sheet 68065

Contact Name	Pierce Laurens
Company/ Hospital Name	Queen Elizabeth The Queen Mother Hospital
Department	EME Department
Direct Phone	01843 234346
General Phone	01843 22 55 44
Opera Account	00003665
Email	pierce.laurens@nhs.net
Order Number	44002074
Date Received	10/Feb/2022
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN34697	1410000	PR0542A12	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	1430309	V1000 Transducer interface cushion.
Qty: 1	1480000	Foetal Heart Simulator V1000 - Service

V1000 has had transducer interface cushion replaced and calibration has been checked
