

Service Repair Sheet SRS68070

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Ashley Coley

Company/ Hospital Name

Conquest Hospital

Department

Position

Direct Phone

03001315151

General Phone

01424 75 52 55

Opera Account

00004831

Email

ashleycoley@nhs.net

Order Number

Date Received

14/Feb/2022

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Quote

Notes 10/Feb/2022 Zoey Teal

10/Feb/2022 Zoey Teal

Sending in a V1000 for repair. Serial number 16755. Customers says I have a Foetal Heart Simulator that is failing to turn on, following over-exposure to ultrasound gel due to a brittle bottle, it is suspected that some of the gel may have ingressed internally.

14/Feb/2022 Robert Connor

Received 1 x V1000 s/n PR0545A11, with 4 x AA battery and grey silicone cover.

Ready For quote

CGreen 14.2.22.

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34736	Foetal Simulator	1410000	PR0545A11	N

14800000 x 1 @ £45
S/N, SRS, SRN

1430309 x 1 @ £0 FOC
SRS, SRN

1430295 x 1 £35
SRS, SRN x 1 £35
SRS, SRN.

UPS x 1 £10.

f: SR568070

EME Department

Quality Management System

East Sussex Healthcare **NHS**
NHS Trust

EQUIPMENT RETURN FORM

OP.7.4.3

TO: SERVICE DEPARTMENT VIAMED, 15 STATION ROAD, CROSS HILLS, KEIGHLEY BD20 7DT		DATE: 10-Feb-22
DESCRIPTION of equipment: V1000 PRO545A11		
Contamination details: Has been exposed to blood, body fluid etc ULTRASOUND GEL ONLY. Has been disinfected/cleaned YES		
Method of Disinfection/cleaning used: CLINELL UNIVERSAL WIPES		
DATE of purchase (approx): 26.02.2009		
REASON FOR RETURN: EXPOSURE TO EXCESS ULTRASOUND GEL NOW NOT POWERING ON.	ACTION REQUIRED: PLEASE PROVIDE AN ESTIMATE FOR REPAIR. AN ORDER CAN BE OBTAINED FROM THE CONTACT DETAILS BELOW WHEN FULL COST IS KNOWN.	
OUR Ref: 640120		
CONTACT: EME DEPARTMENT, ASHLEY COLEY DISTRICT GENERAL HOSPITAL, KINGS DRIVE, EASTBOURNE, EAST SUSSEX. BN21 2UD		
TEL: 0300 131 5151	e-mail: ashleycoley@nhs.net	

Any Person Identifiable Information (PII) that this device may hold is the property of East Sussex Healthcare Trust (ESHT are the Data Controller for the PII on this device). Any access or processing of any PII on this device by you, your company, its servants or agents may breach Data Protection legislation. By being in receipt of this device you are bound by ESHT's confidentiality processes.



Further information regarding PII is available from Esh-tr.infogov@nhs.net

Catherine Spence Required Reading 1 Calls 17 Issues 9 Unread 7 ISO Tasks |

H/W: DL KG Future Active Users

Issues

▼ Search



Select Currency

Pounds | Euros | Usd

Repair SRS68070 Values in Pounds @ £1: £1.00

SRN34736	Foetal Simulator	1410000			
	Parts Required	Reference	Quantity	Unit Cost	Total Line
		1430295	1	£35.00	£35.00
		1430309	1	£15.00	£15.00
		1480000	1	£45.00	£45.00
				Parts Cost	£95.00

FOC with Service
cy 14.2.22

CHECK PRICES BEFORE SENDING QUOTE

Quote Sent