

Service Repair Sheet SRS68064

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Michael Goddard

Addenbrookes Hospital

Clinical Engineering Department

01223 245151

00000860

michael.goddard@addenbrookes.nhs.uk

09/Feb/2022

Robert Connor

Viamed

Quote

Notes 07/Feb/2022 Sophie Lines

07/Feb/2022 Sophie Lines

Sending V1000 for Service/Calibration - aware cost is £45 plus £10 carriage. Serial number: PRO2380A11

09/Feb/2022 Robert Connor

1 x V1000 s/n PRO2380A11, with grey cover and 4 x AA battery

Ready For quote

CGreen 10.2.22.

Repair Complete Signed

CGreen 17.2.22

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34691	Foetal Simulator	1410000	PRO2380A11	N

1480000 x 1 @ £45
S/N, SRS, SRN.

1430309 x 1 FOC
SRS, SRN

UPS x 1 FOC

**Viamed Ltd
15 Station Road
Cross Hills
Keighly
West Yorkshire
BD20 7DT**

Date: 07 February 2022

Dear Sir/Madam

Please find enclosed one V1000 foetal heart simulator (serial number PRO2380A11) which we are returning to you for service and certificate calibration, at a cost of £45 + £10 carriage charge. Our returns reference number is SRS68064. We will raise a purchase order to cover the costs once we have received a quotation from you – I hope this is acceptable.

Please do not hesitate to contact me if you have any questions, and many thanks for your help. The return address for the device is below:

Clinical Engineering (Box 152)
Level 5 X-Ray Block
Addenbrookes Hospital
Hills Road
Cambridge
CB2 2QQ

Many thanks again,

Michael Goddard

Clinical Technologist

Q-Pulse Document No.: CE-FRM-17

Revision: 3

Printed on: 07/02/2022



Innovation and excellence in health and care Addenbrookes Hospital | Rosie Hospital

NIHR – Cambridge Biomedical Research Centre | Academic Health Science Centre – Cambridge University Health Partners

Order Date : 14-02-2022
Order No : **PO-000078095**
Must be quoted on all correspondence.

Deliver To :

(1) CLINICAL ENGINEERING L5
Cambridge University Hospitals NHS Foundation
Trust
Addenbrookes Hospital Hills Rd
Cambridge
CB2 0QQ
GB
Requested delivery date: 13-02-2022

Invoice and Payment Enquiries To

Finance Department : Box 130
Cambridge University Hospitals NHS Foundation Trust
Addenbrookes Hospital Hills Road
Cambridge
CB2 0QQ
GB

All enquiries regarding this order to:

Contact : Emma Cowdrey
Telephone :
Facsimile No. :
Email Address : Emma.Cowdrey@addenbrookes.nhs.uk
Buyer Contact : Mark Tubby
Buyer Email : mark.tubby@addenbrookes.nhs.uk

Supplier**Viamed Ltd**

Internal Department Description:
JW1851 Clinical Engineering Lvl 5 - 1

Conditions

This order is issued in accordance with the Trusts and relevant NHS Terms and Conditions of Contract, available at (http://www.dh.gov.uk/en/Publicationsandstatistics/Publications/PublicationsPolicyAndGuidance/DH_121260).

Delivery charges will not be paid unless specified on the order. If there are any irregularities please contact buyer prior to processing order.

The above order number (in the format PO-xxxxxxx) must be quoted on all delivery notes, invoices and correspondence. Failure to do this may result in delay of payment. N.B. Information may be disclosed under the Freedom of Information Act 2000

Goods will only be received between 8:00 and 16:00 Mon to Fri

EORI Number: GB654910528000 - freight logistics & customs broker: Blacksmith Freight Ltd angela@blacksmithfreight.co.uk +44 (0) 161 499 0804

Line	Goods or Services Required	Quantity	UOM	Contract Ref.	Unit Price	Line Value	VAT
1	x1 manufacturer calibration of test equipment with serial number PRO2380A11, as per quote QVM135385. (1) CLINICAL ENGINEERING L5	1	EACH		£45.00	£45.00	-
2	Postage (1) CLINICAL ENGINEERING L5	1	EACH		£10.00	£10.00	-

Net Total :	£55.00
Carriage :	-
Tax :	-
Total	£55.00

Service Repair Sheet 68064

Contact Name

Michael Goddard

Company/ Hospital Name

Addenbrookes Hospital

Department

Clinical Engineering Department

General Phone

01223 245151

Opera Account

00000860

Email

michael.goddard@addenbrookes.nhs.uk

Order Number

PO-000078095

Date Received

09/Feb/2022

Booked in By

Robert Connor

Repair

Ref

S/N

Equipment Type

SRN34691

1410000

PR02380A11

Foetal Simulator

Complete - Repaired

Time :0 Hour(s)

Parts Replaced

Qty: 1

1480000

Foetal Heart Simulator V1000 - Service

Qty: 1

1430309

V1000 Transducer interface cushion.

V1000 has had transducer interface cushion replaced and calibration has been checked
