

Service Repair Sheet SRS68066

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Phuong Trinh

An An Import Export Co Ltd

+84 987 546 072

CID15101

phukienthaythe@gmail.com

10/Feb/2022

Robert Connor

Viamed

Warranty

Notes 07/Feb/2022 Ryan Swaine

07/Feb/2022 Ryan Swaine

MAX-250E from RVM134243-1:

34x failed on the output specification.

GD11999040

GD11999041

GD11999082

GD11999063

GD11999066

GD11999044

GD11999065

GD11999067

GD11999028

GD11999069

GD11999042

GD11999030

GD11999051

GD11999061

GD11999037

GC12999438

GD11999049

GC12999441

GD11999036

GD11999059

GC12999382

GC12999446

GD11999074

GD11999035

GD11999054

GD11999029

GC12999469

GD11999073

GD11999068

GD11999052

GD11999047

GC12999385

GC12999371

GC12999464

2x (GC12999455 & GC12999379) have a gas leak, may be through case, I will ask for more information.

10/Feb/2022 Robert Connor

Received MAX-250E x 34, s/n GC12999371, GC12999379, GC12999382, GC12999385, GC12999438, GC12999441, GC12999446, GC12999455, GC12999464, GC12999469, GD11999028, GD11999029, GD11999030, GD11999035, GD11999036, GD11999037, GD11999040, GD11999041, GD11999042, GD11999044, GD11999047, GD11999049, GD11999051, GD11999052, GD11999054,

GD11999074, GD11999082.
16/Feb/2022 Catherine Spence
engineers room
16/Feb/2022 Philip Crossley
All failed, some Low Output, some Unstable and Nonlinear
17/Feb/2022 Catherine Spence
Spoken to Steve Nixon to make him aware that there are 34 sensors failed maxtec sensors failed, to process as noemal and he will have a chat with maxtec

Ready For quote _____

Repair Complete Signed _____

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34698	Oxygen Sensor	0110429	GC12999371	Yes
SRN34699	Oxygen Sensor	0110429	GC12999379	Yes
SRN34700	Oxygen Sensor	0110429	GC12999382	Yes
SRN34701	Oxygen Sensor	0110429	GC12999385	Yes
SRN34702	Oxygen Sensor	0110429	GC12999438	Yes
SRN34703	Oxygen Sensor	0110429	GC12999441	Yes
SRN34704	Oxygen Sensor	0110429	GC12999446	Yes
SRN34705	Oxygen Sensor	0110429	GC12999455	Yes
SRN34706	Oxygen Sensor	0110429	GC12999464	Yes
SRN34707	Oxygen Sensor	0110429	GC12999469	Yes
SRN34708	Oxygen Sensor	0110429	GD11999028	Yes
SRN34709	Oxygen Sensor	0110429	GD11999029	Yes
SRN34710	Oxygen Sensor	0110429	GD11999030	Yes
SRN34712	Oxygen Sensor	0110429	GD11999036	Yes
SRN34713	Oxygen Sensor	0110429	GD11999037	Yes
SRN34714	Oxygen Sensor	0110429	GD11999040	Yes
SRN34715	Oxygen Sensor	0110429	GD11999041	Yes
SRN34716	Oxygen Sensor	0110429	GD11999042	Yes
SRN34717	Oxygen Sensor	0110429	GD11999044	Yes
SRN34718	Oxygen Sensor	0110429	GD11999047	Yes
SRN34719	Oxygen Sensor	0110429	GD11999049	Yes
SRN34720	Oxygen Sensor	0110429	GD11999051	Yes
SRN34722	Oxygen Sensor	0110429	GD11999054	Yes
SRN34723	Oxygen Sensor	0110429	GD11999059	Yes
SRN34724	Oxygen Sensor	0110429	GD11999061	Yes
SRN34725	Oxygen Sensor	0110429	GD11999063	Yes
SRN34726	Oxygen Sensor	0110429	GD11999065	Yes
SRN34727	Oxygen Sensor	0110429	GD11999066	Yes
SRN34728	Oxygen Sensor	0110429	GD11999067	Yes
SRN34729	Oxygen Sensor	0110429	GD11999068	Yes
SRN34730	Oxygen Sensor	0110429	GD11999069	Yes
SRN34731	Oxygen Sensor	0110429	GD11999073	Yes
SRN34732	Oxygen Sensor	0110429	GD11999074	Yes
SRN34733	Oxygen Sensor	0110429	GD11999082	Yes