

Invoice Correction 5259011430**ORIGINAL**

Invoice Correction Date: 02.02.2022

Honeywell Healthcare Solutions GmbH

Alter Holzhafen 18

Purchase Order No: RTS3100833

23966 Wismar

Germany

Bill to: 1797471Vandagraph Limited
15 Station Road
Cross Hills, West Yorkshire
BD20 7DT
KEIGHLEY
United Kingdom**Sold to:** 1797448Vandagraph Sensor Technologies Ltd.
15 Station Road
Cross Hills, West Yorkshire
KEIGHLEY
BD20 7DT
United Kingdom**Order Information:**Credit Memo Request: 63791669
Sales Office: 2170 DE SPS Wismar
Incoterms: FCA Origin
Payment Terms: Net 30 Days After Invoice Date**Ship to:** 1797448Vandagraph Sensor Technologies Ltd.
15 Station Road
Cross Hills, West Yorkshire
KEIGHLEY
BD20 7DT
United Kingdom

ITEM NO.	Material No.	QTY	UOM	UNIT PRICE	EXT. PRICE	VAT %
000010	E1002734 Oxygen Sensor OOD103 VAN 8010008 Customer Part Nbr: 8010008 credit note for SN: 110475	1	EA	18,66 ✓	-18,66	0%
000020	E1002749 Oxygen Sensor OOD103 VAN 8010026 Customer Part Nbr: 8010026 credit note for SN: 100233	1	EA	18,66 ✓	-18,66	0%

Totals

Inflation Surcharge 8,00%	-2,98 EUR
Total Before VAT	-40,30 EUR
VAT-Total	0,00 EUR
Total Amount	-40,30 EUR

Tax Information

Tax Exempt Reason: Tax free export - Goods

Duns #:

Bank Details: No Payment Due

Account Name:

Account No:

Bank Key:

Swift Code:

IBAN:

Payment method:

Registered No: Schwerin HRB 1838

VAT Reg No: DE137437213

Federal ID #:

Registered Address: Honeywell Healthcare Solutions GmbH, Alter Holzhafen 18, Wismar, 23966, Germany

Credit Analyst: Xenia Musat

Sales Rep: Stefan Dreyer

Customer Care Rep: Jessica Wagner +49 3841 360 206

Honeywell Healthcare Solutions GmbH

WEEE Reg.No.: 57798680
Managing Director: Marcus OstländerInternet: www.envitec.com
email: info@envitec.com

Delivery exclusive to our General Terms (February 2016). We retain title to our goods and deliveries until all the payment from the sales contract have been received. Your order will be executed exclusively to our terms and conditions. These can be sent by separate request or viewed at www.envitec.com

PRÜFBERICHT - TEST REPORT Sauerstoffsensoren MED / Oxygen Sensors Medical Use

Vorgangsnummer/ Process Number	47867391	Benachrichtigungsnummer/ Notification Number	300247392
-----------------------------------	----------	---	-----------



Kunde / Customer <i>Vanda graph</i>	Kunden - Auftragsnummer / Customer Order No. RTS3100833
Auftragsdatum / Date of Order 20.12.2021	
Lief.an code/ Delivery Code	Aufgenommen von / Received by <i>KG</i>
Beanstandung / Objection	
<i>zu langsam</i>	

Artikelnummer / Part No.	Herstell. Datum / Date of Manufact.	Code + Seriennummer / Code + Serial No.	Garantie bis / End of Warranty Date	Beschreibung / Description
E1002734	<i>08/21</i>	110475	<i>10/22</i>	Oxygen Sensor OOD103 VAN 8010008

Prüfkriterien unter Standardbedingungen / Test Parameters under Standard Conditions

- | | | |
|---|---|------------------|
| 1 | Ausgangssignal an Luft / Output Signal in Atmosphere | [20,95 % vol.O2] |
| 2 | Linearitätsfehler, Meßfehler / Linearity Error, Error of Measurment | [100 % vol.O2] |
| 3 | Nullsignal / Zero Signal | [100 % vol.N2] |

Messprotokoll / Test Report

Prüfkriterien / Test Parameters	Spezifiziert / Specified	Gemessen / Measured	
		Signal1	Signal1
1 [mV]	<i>11.2</i>	<i>11.4</i>	
2 [%]	<i>6200</i>	<i>27</i>	
3 [µV]	<i>43</i>	<i>0.4</i>	

Prüfbericht / Checklist

Fehler / Defect:	O2-S001..Luftwert außerh.Toleranz / Signal Out Of Tolerance	
	O2-S002..Signal Instabil / Signal Not Stable	
	O2-S003..Offset Außerh. Toleranz / Offset Out Of Tolerance	
	O2-S004..Signal 0mV / Signal 0mV	
	O2-S005..Gehäuse Defekt / Housing Defective	
	O2-S006..Elektrolytaustritt / Electrolyte Leakage	
	O2-S007..Sonstiges / Others	
Bemerkungen / Remark	kein Fehler festgestellt, Sensor innerhalb Spezifikation / No failure found, Sensor within specifications	☒
	außerhalb der Garantie / out of warranty	

Anerkannt / Accepted ☐	Nicht anerkannt / Not Accepted ☒
---	--

Austausch / Replacement	Rücksendung / Return
VB-Nr. / VB No	Lieferdatum / Delivery Date
Seriennummer / Serial No.:	Geprüft / Inspected: <i>Bjll</i>
Lieferdatum / Delivery Date Bearbeitet / Processed <i>26.01.22</i>	