

Service Repair Sheet SRS68078

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Lee Morris

Spire Wellesley Hospital

Biomedical Equipment Manager

01702462944

01702 447913

00000151

lee.morris@spirehealthcare.com

21/Feb/2022

Robert Connor

Viamed

Quote

Notes Sending a DB3 unit back for repair. Emailed attached for reference.

17/Feb/2022 Sarah Walton

The black button is not making positive click when turned â€" it rotates full 360â€™

21/Feb/2022 Robert Connor

Received 1 x DB3 s/n M0004930, in hard case with battery and patient lead.

Ready For quote

*C.Green 25.2.2022*

Repair Complete Signed

---

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34742	Microstim DB3	2510000	M0004930	N

*2540010 x 1 @ £45*

*S/N, SRS, SRN*

*UPS x 1 @ £10*

Declaration of decontamination status prior to the inspection,  
servicing or return of medical and laboratory equipment.

GETINGE \*

To: VIAIR

14 STATION RD

CROSS HILLS KEIGHLEY

WEST YORKSHIRE BD 20 7 DT

Ref  
SRS 68078

REPAIR

Retain this document and  
quote serial number if  
you have a query.

Serial No.:  
300159

Tick box A if applicable, otherwise complete all items of part B providing further information as requested or appropriate.

**A**



This equipment/ item has not been used in any invasive procedure or been in contact with blood, other body fluids, respired gases or pathological samples. It has been cleaned in preparation for inspection, servicing, repair or transportation.

**B**

1. Has this equipment/ item been exposed internally or externally to hazardous materials as indicated below?

\* Complete as appropriate.

\* **YES/NO**

Blood, body fluids, respired gases, pathological samples:

\* **YES/NO**

Other bio-hazards:

\* **YES/NO**

Chemicals or substances hazardous to health:

\* **YES/NO**

Other hazards:

2. Has the equipment/ item been cleaned and decontaminated?

\* **YES/NO**

Indicate the methods and materials used:  
If the equipment/ item could not be decontaminated,  
please indicate why:

3. Has the equipment/ item been suitably prepared to ensure safe handling/ transportation?

\* **YES/NO**

I declare that I have taken all reasonable steps to ensure the accuracy of the above information in accordance with HSG (93)26.

Authorized signature: [Signature]

Date: 17/2/22

Print name: LEE MORRU

Position: BIO MEDICAL EQUIP MANAGER

Department: THEATRES

Hospital: SPRUE WELLSLEY

SOUTHERN

Provide further details here:

REPAIR

SRS 68078


This item must be treated as  
decontaminated only.

UPS x 1 @ £10

Catherine Spence Required Reading 1 Calls 36 Issues 27 Unread 17 ISO Tasks |

H/W: HL SH SN Future Active Users

Search



### Select Currency

Pounds| Euros| Usd

Repair SRS68078 Values in Pounds @ 1: 1.00

SRN34742	Microstim DB3	2510000			
	Parts Required	Reference	Quantity	Unit Cost	Total Line
		2540010	1	45.00	45.00
				Parts Cost	45.00

CHECK PRICES BEFORE SENDING QUOTE

Quote Sent