

Service Repair Sheet SRS68079

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Jamie Bennett

Company/ Hospital Name

Chesterfield Royal Hospital

Department

Position

Direct Phone

General Phone

01246 277 271

Opera Account

00001170

Email

jamie.bennett@nhs.net

Order Number

80054485

Date Received

21/Feb/2022

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Repair

Notes PO has now been received for this repair which I have attached

21/Feb/2022 Sarah Walton

SN: H0002417 Faulty DB3 unit being sent back

21/Feb/2022 Robert Connor

Received 1 x DB3 s/n H0002417, with battery and patient lead.

Purchase order
received
cy 25.2.22

Ready For quote

Repair Complete Signed

CONNOR 28.2.22

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34741	Microstim DB3	2510000	H0002417	N

2540010 x 1 @ £45
S/N, SRS, SRN

UPS x 1 @ £10

Purchase Order

Derbyshire Support
and Facilities Services

VIAMED LTD
15 STATION ROAD
CROSSHILLS
KEIGHLEY
WEST YORKSHIRE
BD20 7DT

Order number 80054485
Order date 18/02/2022
Delivery attention
Payment terms 30 Days
Your ref Viamed FPR

Deliver to:
Chesterfield Royal Hospital
Receipts & Distribution
Top Road, Calow, Chesterfield, S44 5BL

Tel No. 01246 277271
Fax No. 01246 512610
Email: CRHFT.Procurement@nhs.net

Invoice to:
DSFS Limited
c/o Chesterfield Royal Hospital
Financial Services Department, Top Road, Calow,
Chesterfield, Derbyshire, S44 5BL

Email: crhft.dsfs.pdfinvoices@nhs.net

Our opening hours are 9am - 4pm
All correspondence must contain our order number

Page 1 of 1

Article	Description	Qty	Unit	Unit Price	Amount in GBP	Delivery Date
1115961	Clinical Engineering Spare Parts [Please see details] Fixed Price Repair of Viamed Microstim DB3 Stimulator - SN: H0002417 £45.00	1.00	EA	45.00	45.00	25/02/2022
1115961	Clinical Engineering Spare Parts [Please see details] Carriage: £10	1.00	EA	10.00	10.00	25/02/2022

Subject to NHS Model terms and conditions
for purchase of goods and services, unless
otherwise stated.

Total Net 55.00
Total VAT 11.00
Total in GBP 66.00

Service Repair Sheet 68079

Contact Name

Jamie Bennett

Company/ Hospital Name

Chesterfield Royal Hospital

General Phone

01246 277 271

Opera Account

00001170

Email

jamie.bennett@nhs.net

Order Number

80054485

Date Received

21/Feb/2022

Booked in By

Robert Connor

Repair

Ref

S/N

Equipment Type

SRN34741

2510000

H0002417

Microstim DB3

Complete - Repaired

Time :0 Hour(s)

Parts Replaced

Qty: 1 2540010

The Microstim DB3 has been repaired and the patient lead has been replaced.
