


Catherine Spence [Required Reading](#) **1 Calls** 35 Issues **24 Unread** 18 ISO Tasks | [H/W:](#)
[Future](#) [Active Users](#)
 Search

XHALE S.A.

CID9388 Opera **00006401** Type: **Company Distributor**
XHALE S.A. Evagelistrias 3, Kallithea, Athens, Greece, , 17671, Greece
 Tel **+30 693 660 9908** [Log Call](#)
 Proforma VAT EL800753522



[Serial number lists for opera](#)

Company / Hospital	Xhale
Contact Name	Apostolos Atsalakis
Department	
Position	
Company Phone	00302109531294
Direct Phone	
Opera Account	00006401
Viamed ORD Number	
Email	apostolos@endomedical.co
Order Number	
Type	Warranty

SRS68036 Warranty Repair

Xhale ,Apostolos Atsalakis

SRS68036 Notes:

05/Jan/2022 Catherine Spence

05/Jan/2022 Catherine Spence

46 UFO sensors received, out of these 7 of them are from RVM134513-1 from 16/12/2021 which was a replacement order(check with Steve Nixon about these.

(S/N 689220,689223-689224,289228,689230,689233,689236)

05/Jan/2022 Catherine Spence

Received on DHL AWB7354542300

12/Jan/2022 Philip Crossley

They are in my room. I cannot test these.

13/Jan/2022 Catherine Spence

39 units do not have the wings on the sensors

21/Jan/2022 Catherine Spence

Steve Nixon aware these are here.

25/Jan/2022 Catherine Spence

Test unit has been made and this repair is now in testing

05/Feb/2022 Catherine Spence

tested see report attached to this srs

09/Feb/2022 Catherine Spence

need to remove the duplicate line for S/N 684376 error

09/Feb/2022 Catherine Spence

waiting on new stock to arrive , also waiting on report from Teledyne on other sensors that have already been returned from other SRS's, Steve Nixon to advise

16/Feb/2022 Catherine Spence

engineers room, waiting for replacements

02/Mar/2022 Catherine Spence

noticed on second checking that S/N 672028 has a broken clip

03/Mar/2022 Robert Connor

Removed duplicate serial number 684376

03 /03 / 2022 Catherine Spence

Move to new srs due to broken clip

Move SRN34512 From SRS68036 To SRS68095

03 /03 / 2022 Catherine Spence

Moved to another SRS waiting for replacement

Move SRN34528 From SRS68036 To SRS68096

Move SRN34530 From SRS68036 To SRS68096

Move SRN34531 From SRS68036 To SRS68096

Move SRN34532 From SRS68036 To SRS68096

Move SRN34533 From SRS68036 To SRS68096

Move SRN34534 From SRS68036 To SRS68096

Move SRN34536 From SRS68036 To SRS68096

Move SRN34537 From SRS68036 To SRS68096

Move SRN34538 From SRS68036 To SRS68096

Move SRN34539 From SRS68036 To SRS68096

Move SRN34540 From SRS68036 To SRS68096

Move SRN34541 From SRS68036 To SRS68096

Move SRN34542 From SRS68036 To SRS68096

Move SRN34543 From SRS68036 To SRS68096

Add Notes to SRS

All items for repair

[Add Blank Line For new item linked to repair](#)

SRN	Equipment	Reference	Serialnumber Barcode	Serialnumber	Description	Customer Warranty New	Customer Warranty Repair	Our Supplier Warranty	Status	History
SRN34494	Oxygen Sensor	0110081	1851667	667938	Oxygen sensor, UFO 130-2	Yes	No	☐ Check	Return To Supplier	History
SRN34494 Customer Notes: Unstable output										
SRN34495	Oxygen Sensor	0110081	1887469	672004	Oxygen sensor, UFO 130-2	Yes	No	☐ Check	Return To Supplier	History
SRN34495 Customer Notes: Unstable output										
SRN34498	Oxygen Sensor	0110081	1887470	672005	Oxygen sensor, UFO 130-2	Yes	No	☐ Check	Return To Supplier	History
SRN34498 Customer Notes: Unstable output										
SRN34499	Oxygen Sensor	0110081	1887472	672007	Oxygen sensor, UFO 130-2	Yes	No	☐ Check	Return To Supplier	History
SRN34499 Customer Notes: Unstable output										
SRN34500	Oxygen Sensor	0110081	1887473	672008	Oxygen sensor, UFO 130-2	Yes	No	☐ Check	Return To Supplier	History
SRN34500 Customer Notes: Unstable output										
SRN34501	Oxygen Sensor	0110081	1887475	672010	Oxygen sensor, UFO 130-2	Yes	No	☐ Check	Return To Supplier	History
SRN34501 Customer Notes: Unstable output										
SRN34502	Oxygen Sensor	0110081	1887476	672011	Oxygen sensor, UFO 130-2	Yes	No	☐ Check	Return To Supplier	History
SRN34502 Customer Notes: Unstable output										
SRN34503	Oxygen Sensor	0110081	1887477	672012	Oxygen sensor,	Yes	No	☐ Check	Return To	History

					UFO 130-2				Supplier	
SRN34503 Customer Notes: Unstable output										
SRN34504	Oxygen Sensor	0110081	1887478	672013	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34504 Customer Notes: Unstable output										
SRN34505	Oxygen Sensor	0110081	1887481	672016	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34505 Customer Notes: Unstable output										
SRN34506	Oxygen Sensor	0110081	1887483	672018	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34506 Customer Notes: Unstable output										
SRN34507	Oxygen Sensor	0110081	1887484	672019	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34507 Customer Notes: Unstable output										
SRN34508	Oxygen Sensor	0110081	1887486	672021	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34508 Customer Notes: Unstable output										
SRN34509	Oxygen Sensor	0110081	1887489	672024	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34509 Customer Notes: Unstable output										
SRN34510	Oxygen Sensor	0110081	1887490	672025	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34510 Customer Notes: Unstable output										
SRN34511	Oxygen Sensor	0110081	1887490	672026	Oxygen sensor, UFO 130-2	Yes Change	No Change	Check	Return To Supplier	History
SRN34511 Customer Notes: Unstable output										
SRN34513	Oxygen Sensor	0110081	1887496	672031	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34513 Customer Notes: Unstable output										
SRN34514	Oxygen Sensor	0110081	1887504	672039	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34514 Customer Notes: Unstable output										
SRN34515	Oxygen Sensor	0110081	1887505	672040	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34515 Customer Notes: Unstable output										
SRN34516	Oxygen Sensor	0110081	1887506	672041	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34516 Customer Notes: Unstable output										
SRN34517	Oxygen Sensor	0110081	1887514	681626	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34517 Customer Notes: Unstable output										
SRN34518	Oxygen Sensor	0110081	1887523	681635	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34518 Customer Notes: Unstable output										
SRN34519	Oxygen Sensor	0110081	1887545	681657	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34519 Customer Notes: Unstable output										

04/03/2022, 08:00

Repair System

SRN34520	Oxygen Sensor	0110081	1890978	684319	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34520 Customer Notes: Unstable output										
SRN34521	Oxygen Sensor	0110081	1890979	684320	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34521 Customer Notes: Unstable output										
SRN34522	Oxygen Sensor	0110081	1890986	684327	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34522 Customer Notes: Unstable output										
SRN34523	Oxygen Sensor	0110081	1890992	684333	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34523 Customer Notes: Unstable output										
SRN34524	Oxygen Sensor	0110081	1891005	684346	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34524 Customer Notes: Unstable output										
SRN34525	Oxygen Sensor	0110081	1891009	684350	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34525 Customer Notes: Unstable output										
SRN34526	Oxygen Sensor	0110081	1891021	684362	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34526 Customer Notes: Unstable output										
SRN34527	Oxygen Sensor	0110081	1891029	684370	Oxygen sensor, UFO 130-2	Yes	No	Check	Return To Supplier	History
SRN34527 Customer Notes: Unstable output										

SRN34511 Oxygen Sensor 0110081 672026 Oxygen sensor, UFO 130-2
[Global fill fields across multiple 0110081's records on SRS](#)

[Create Issue - Use to link files and or photos to the SRS](#)
Internal Engineer Notes against SRN34511

03/Mar/2022 Catherine Spence
Replaced in advance of sensors being evaluated for route cause analysis. Replaced with S/N:778845

Rising output during use >15 mV

(add separately to completing a repair)

Quote Item Against SRN34511 Request Item Against SRN34511 Add Labour Item SRN34511						
Qty	Item	Description	By	Date Requested	Status	
1	0110081	Oxygen sensor UFO 130-2	Catherine Spence	03/Mar/2022	completed	Remove
Refresh List						

[Fill in Q.A. Form For Oxygen Sensor 672026](#)

Qid	Question	Field	Updated Date
3098	Pass	☐	
3099	Output mV		Save
3103	Air signal out of tolerance	☒ low ☐	

3104 Air signal out of tolerance	high	☐
3105 Unstable output signal		☐
3106 Zero offset out of tolerance		☐
3107 No output		☐
3108 Housing defective		☐
3109 Electrolyte leakage		☐
3110 Linearity out of tolerance		☐
3111 Gas Leak Test Fail		☐
3112 ShortNotes		Save

Notes Customer will read

Replaced in advance of sensors being evaluated for route cause analysis. Replaced with S/N:778845

Complete Repair

Chargable Time	Actual Repair Time	Fault Code	Status	
Zero Seconds ▼	Zero Seconds ▼	Rising output during use >15 mV ▼	Return To Supplier ▼	Submit

Quote Preview

Repair SRS68036 Values in Pounds @ £1: £1.00

Parts Cost	£0.00
---	------------------------------------