

Service Repair Sheet SRS68078

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Lee Morris

Spire Wellesley Hospital

Biomedical Equipment Manager

01702462944

01702 447913

00000151

lee.morris@spirehealthcare.com

21/Feb/2022

Robert Connor

Viamed

Quote

Notes Sending a DB3 unit back for repair. Emailed attached for reference.

17/Feb/2022 Sarah Walton

The black button is not making positive click when turned â€" it rotates full 360â€™

21/Feb/2022 Robert Connor

Received 1 x DB3 s/n M0004930, in hard case with battery and patient lead.

Ready For quote

CGreen 25.2.2022

Repair Complete Signed

CGreen 4-3-22

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34742	Microstim DB3	2510000	M0004930	N

2540010 x 1 @ £45
S/N, SRS, SRN

UPS x 1 @ £10

PURCHASE ORDER

Order No.: 4503580980
To be quoted on all documents

Order date: 28/02/22
Date delivery required: 28/02/22
Invoices should be sent to:

SHL Spire Healthcare Ltd.
25-41 Crown St., Reading RG1 2PQ

Fax: 01702 601049
Telephone: 01702 447913

Delivery Address:
WLS Other
Spire Wellesley Hospital
Eastern Avenue
Southend-on-Sea, Essex
SS2 4XH

Supplier:
AMED LIMITED
Station Rd
Bosshills
Highley
East Yorkshire
YO20 7DT
Fax: 01535 635 582

use supply the goods / services listed below
subject to the terms and conditions set out on the last page)

Line	Spire Code	Quantity and Order Unit	Description	Catalogue number	Price Per Unit	VAT	Total price
10		1 EA	Microstim DB3 Supramaximal Nerve Stimula	QVM135664	45.00	20.00% 9.00	45.00
Sub total:							45.00
VAT:							9.00
Total:							54.00

For and on behalf of Spire Healthcare Ltd.
Registered Offices: 3 Dorset Rise, London EC4Y 8EN.
Registered in England and Wales no: 1522532
(not for invoice enquiries or addressing of invoices)