

Service Repair Sheet SRS68078

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Lee Morris

Spire Wellesley Hospital

Biomedical Equipment Manager

01702462944

01702 447913

00000151

lee.morris@spirehealthcare.com

21/Feb/2022

Robert Connor

Viamed

Quote

Notes Sending a DB3 unit back for repair. Emailed attached for reference.

17/Feb/2022 Sarah Walton

The black button is not making positive click when turned "it rotates full 360°"

21/Feb/2022 Robert Connor

Received 1 x DB3 s/n M0004930, in hard case with battery and patient lead.

Ready For quote

Cgreen 25.2.2022

Repair Complete Signed

Cgreen 4-3-22

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34742	Microstim DB3	2510000	M0004930	N

2540010 x 1 @ £45

S/N, SRS, SRN

2530012 x 1 FOC

SRS, SRN

2520000 x 1 FOC

SRS, SRN

UPS x 1 @ £10

cost  
included in the fixed  
Part of the charge - (g 4-3-22)