

Service Repair Sheet 68062
 Contact Name William Ewles
 Company/ Hospital Name Medway Maritime Hospital
 Department Equipment Services
 Direct Phone 01634976264
 General Phone 01634 83 00 00
 Opera Account 00001770
 Email william.ewles@nhs.net
 Order Number 220019824
 Date Received 08/Feb/2022
 Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN34670	0310030	E6	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	0380000
Qty: 1	0330217
Qty: 1	0330201
Qty: 1	0330210
Qty: 1	0330190
Qty: 1	0330191

Tom Thumb E6 has been serviced and rebuilt with new block, precision valve, clamp, spacer and oxygen hose.

Repair	Ref	S/N	Equipment Type	
SRN34671	0310030	E9	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	0380000
Qty: 1	0330217
Qty: 1	0330201
Qty: 1	0330190
Qty: 1	0330191
Qty: 1	0330210

Tom Thumb E9 has been serviced and rebuilt with new block, precision valve, clamp, spacer and oxygen hose.

Repair	Ref	S/N	Equipment Type	
SRN34672	0310030	E12	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	0380000
Qty: 1	0330217
Qty: 1	0330201
Qty: 1	0330190
Qty: 1	0330191
Qty: 1	0330203

Tom Thumb E12 has been serviced and rebuilt with new block, pressure gauge, clamp, spacer and oxygen hose.

Service Repair Sheet 68062
 Contact Name William Ewles
 Company/ Hospital Name Medway Maritime Hospital
 Department Equipment Services
 Direct Phone 01634976264
 General Phone 01634 83 00 00
 Opera Account 00001770
 Email william.ewles@nhs.net
 Order Number 220019824
 Date Received 08/Feb/2022
 Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN34673	0310030	E14	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	0380000
Qty: 1	0330217
Qty: 1	0330201
Qty: 1	0330210
Qty: 1	0330190
Qty: 1	0330191

Tom Thumb E14 has been serviced and rebuilt with new block, precision valve, clamp, spacer and oxygen hose.

Repair	Ref	S/N	Equipment Type	
SRN34674	0310030	VMA9113	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	0380000
Qty: 1	0330217

Tom Thumb TT480 VMA9113 has been stripped and cleaned then reassembled with new O rings and new oxygen hose. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	
SRN34675	0310030	VMA9171	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	0380000
Qty: 1	0330217

Tom Thumb TT480 VMA9171 has been stripped and cleaned then reassembled with new O rings and new oxygen hose. The two valves have been checked and reset before QA.

Service Repair Sheet 68062

Contact Name	William Ewles
Company/ Hospital Name	Medway Maritime Hospital
Department	Equipment Services
Direct Phone	01634976264
General Phone	01634 83 00 00
Opera Account	00001770
Email	william.ewles@nhs.net
Order Number	220019824
Date Received	08/Feb/2022
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN34676	0310030	VMA9175	Tom Thumb	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	0380000		
	Qty: 1	0330217		

Tom Thumb TT480 VMA9175 has been stripped and cleaned then reassembled with new O rings and new oxygen hose. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	
SRN34677	0310030	VMA9176	Tom Thumb	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	0380000		
	Qty: 1	0330217		
	Qty: 1	0330236		

Tom Thumb TT480 VMA9176 has been stripped and cleaned then reassembled with new O rings, gauge face and oxygen hose. The two valves have been checked and reset before QA.
