

Invoice Correction 5259313089

ORIGINAL

Invoice Correction Date: 04.03.2022

Honeywell Healthcare Solutions GmbH

Alter Holzhafen 18

Purchase Order No: RTS1100846

23966 Wismar

Germany

Bill to: 1797550

Viamed Ltd. Cross Hills
15 Station Road
KEIGHLEY
BD20 7DT
United Kingdom

Sold to: 1797550

Viamed Ltd. Cross Hills
15 Station Road
KEIGHLEY
BD20 7DT
United Kingdom

Order Information:

Credit Memo Request: 63813792
Sales Office: 2170 DE SPS Wismar
Incoterms: FCA Origin
Payment Terms: Net 30 Days After Invoice Date

Ship to: 1797550

Viamed Ltd. Cross Hills
15 Station Road
KEIGHLEY
BD20 7DT
United Kingdom

ITEM NO.	Material No.	QTY	UOM	UNIT PRICE	EXT. PRICE	VAT %
000010	E1002718 Oxygen Sensor OOA101-V VIA Customer Part Nbr: 0110182 credit note for SN:103942; 104654	2	EA	12.90	-25.80	0%

Little Low
Should be €13.37.

Totals

Total Before VAT -25.80 EUR
VAT-Total 0.00 EUR
Total Amount -25.80 EUR

Tax Information

Tax Exempt Reason: Tax free export - Goods

Integrity and Compliance at Honeywell

Honeywell places the highest value on the integrity of its business operations. We expect our employees and business partners to comply with our Code of Business Conduct and all applicable laws and regulations. If you become aware of illegal or improper conduct by a Honeywell employee please contact your Honeywell relationship manager. You can also report misconduct directly to us either via e-mail to our Integrity and Compliance Office at access.integrity.helpline@honeywell.com or via our Access helpline accessible worldwide at 800-237-5982.

Duns #:

Bank Details: No Payment Due

Account Name:

Account No:

Bank Key:

Swift Code:

IBAN:

Payment method:

Registered No: Schwerin HRB 1838

VAT Reg No: DE137437213

Federal ID #:

Registered Address: Honeywell Healthcare Solutions GmbH, Alter Holzhafen 18, Wismar, 23966, Germany

Credit Analyst: Xenia Musat

Sales Rep: Stefan Dreyer

Customer Care Rep: Jessica Wagner + 49 3841 360 206

Honeywell Healthcare Solutions GmbH

WEEE Reg.No.: 57798680

Internet: www.envitec.com

Managing Director: Marcus Ostländer

email: info@envitec.com

Delivery exclusive to our General Terms (February 2016). We retain title to our goods and deliveries until all the payment from the sales contract have been received. Your order will be executed exclusively to our terms and conditions. These can be sent by separate request or viewed at www.envitec.com

PRÜFBERICHT - TEST REPORT

ENVITEC
by Honeywell


Sauerstoffsensoren/Oxygen Sensors OOA, OOI, OOD

Kunde/ Customer	Via med	Retour/ RMA- No.	Kunde/ Customer	RTS 1100 846	Envitec	800 248 218
Artikel/ Article	00A 101 - V	Anzahl/ Quantity	2		Eing. Datum/ Inc. Date	15.04.22
Bearbeitet/ processed	Bpm	Prüfdatum/ Insp. Date	16.02.22		Geprüft/ Inspected	Bpm

Produkt/ Identification			Messwert/ Test Result			Bewertung/ Evaluation		
Nr.	Seriennummer/ Serial no.	Lieferlos/ Manufact. code	Luftwert/ Air signal @ 20,9% O ₂ [mV]	Nullsignal/ Offset @ 0% O ₂ [µV]	Linearität/ Linearity [%]	Bemerkung/ Remark	außerhalb Garantie/ out of warranty	kein Fehler festgestellt/ no failure found
1	103942	03121	11,5	24	i.o.			x
2	104654	11121	11,6	21	i.o.	Drucktest		07
3								
4								
5								
6								
7								
8								
9								
10								
11								
12								
13								
14								
15								
16								
17								
18								
19								
20								

Fehler O2-S001: Luftwert außerhalb Toleranz/ Air signal out of tolerance
 Fehler O2-S003: Offset außerhalb Toleranz/ Offset out of tolerance
 Fehler O2-S005: Gehäuse defekt/ Housing defective
 Fehler O2-S007: Sonstiges (unter Angabe in Bemerkung) / Others (see remark)

Fehler O2-S002: Signal instabil / Signal not stable
 Fehler O2-S004: Signal 0mV / Signal 0mV
 Fehler O2-S006: Elektrolyt Austritt / Electrolyte leakage

**BOX:846**

Manufacturer RMA : 300249218

☒ Repaired ☐ No Fault Found

Returned Barcode

Tick All Replaced

Stock Ref	Sup Ref	B/C	S/N	Fault	Replaced	No Fault Found	Repaired	Credited	Out Warranty	Notes
0110182		1718181	V103942	Rising output during use >15 mV Country of Origin: Germany Customs value only: 5.00 Euro	☐	☐	☐	☒	☐	
0110182		1890678	V104654	Rising output during use >15 mV Country of Origin: Germany Customs value only: 5.00 Euro	☐	☐	☐	☒	☐	

Update