

Invoice Correction 5259219279

ORIGINAL

Invoice Correction Date: 24.02.2022

Honeywell Healthcare Solutions GmbH

Alter Holzhafen 18

Purchase Order No: RTS1100835

23966 Wismar

Germany

Bill to: 1797550

Viamed Ltd. Cross Hills
15 Station Road
KEIGHLEY
BD20 7DT
United Kingdom

Sold to: 1797550

Viamed Ltd. Cross Hills
15 Station Road
KEIGHLEY
BD20 7DT
United Kingdom

Order Information:

Credit Memo Request: 63807751
Sales Office: 2170 DE SPS Wismar
Incoterms: FCA Origin
Payment Terms: Net 30 Days After Invoice Date

Ship to: 1797550

Viamed Ltd. Cross Hills
15 Station Road
KEIGHLEY
BD20 7DT
United Kingdom

ITEM NO.	Material No.	QTY	UOM	UNIT PRICE	EXT. PRICE	VAT %
000010	E1002780 Oxygen Sensor OOM112 VIA Customer Part Nbr: 0110043 credit note for SN: 108087; 108099	2	BOX	46.00	-92.00	0%

Totals

Inflation Surcharge 8.00%	-7.36 EUR
Total Before VAT	-99.36 EUR
VAT-Total	0.00 EUR
Total Amount	-99.36 EUR

Tax Information

Tax Exempt Reason: Tax free export - Goods

Integrity and Compliance at Honeywell

Honeywell places the highest value on the integrity of its business operations. We expect our employees and business partners to comply with our Code of Business Conduct and all applicable laws and regulations. If you become aware of illegal or improper conduct by a Honeywell employee please contact your Honeywell relationship manager. You can also report misconduct directly to us either via e-mail to our Integrity and Compliance Office at access.integrity.helpline@honeywell.com or via our Access helpline accessible worldwide at 800-237-5982.

Duns #:

Bank Details: No Payment Due

Account Name:

Account No:

Bank Key:

Swift Code:

IBAN:

Payment method:

Registered No: Schwerin HRB 1838

VAT Reg No: DE137437213

Federal ID #:

Registered Address: Honeywell Healthcare Solutions GmbH, Alter Holzhafen 18, Wismar, 23966, Germany

Credit Analyst: Xenia Musat

Sales Rep: Stefan Dreyer

Customer Care Rep: Jessica Wagner +49 3841 360 206

Honeywell Healthcare Solutions GmbH

WEEE Reg.No.: 57798680

Internet: www.envitec.com

Managing Director: Marcus Ostländer

email: info@envitec.com

Delivery exclusive to our General Terms (February 2016). We retain title to our goods and deliveries until all the payment from the sales contract have been received. Your order will be executed exclusively to our terms and conditions. These can be sent by separate request or viewed at www.envitec.com

PRÜFBERICHT - TEST REPORT

Sauerstoffsensoren MED / Oxygen Sensors Medical Use

Vorgangsnummer/ Process Number	47860587	Benachrichtigungsnummer/ Notification Number	300247850
-----------------------------------	----------	---	-----------



Kunde / Customer <i>Via med</i>	Kunden - Auftragsnummer / Customer Order No.	RTS1100835
------------------------------------	---	------------

Auftragsdatum / Date of Order	05.01.2022
-------------------------------	------------

Lief.an code/ Delivery Code	Aufgenommen von / Received by <i>VG</i>
--------------------------------	--

Beanstandung / Objection

instabil

Artikelnummer / Part No.	Herstell. Datum / Date of Manufact.	Code + Seriennummer / Code + Serial No.	Garantie bis / End of Warranty Date	Beschreibung / Description
E1002780	06/21	108087	08/22	Oxygen Sensor OOM112 VIA

Prüfkriterien unter Standardbedingungen / Test Parameters under Standard Conditions

- | | | |
|---|---|------------------|
| 1 | Ausgangssignal an Luft / Output Signal in Atmosphere | [20,95 % vol.O2] |
| 2 | Linearitätsfehler, Meßfehler / Linearity Error, Error of Measurment | [100 % vol.O2] |
| 3 | Nullsignal / Zero Signal | [100 % vol.N2] |

Messprotokoll / Test Report

Prüfkriterien / Test Parameters	Spezifiziert / Specified	Gemessen / Measured	
		Signal1	Signal1
1 [mV]	25-50	33,1	
2 [%]	2400	7400	
3 [µV]	43	43	

Prüfbericht / Checklist

Fehler / Defect:	O2-S001..Luftwert außerh.Toleranz / Signal Out Of Tolerance	
	O2-S002..Signal Instabil / Signal Not Stable	
	O2-S003..Offset Außerh. Toleranz / Offset Out Of Tolerance	
	O2-S004..Signal 0mV / Signal 0mV	
	O2-S005..Gehäuse Defekt / Housing Defective	
	O2-S006..Elektrolytaustritt / Electrolyte Leakage	
	O2-S007..Sonstiges / Others	
Bemerkungen / Remark	kein Fehler festgestellt, Sensor innerhalb Spezifikation / No failure found, Sensor within specifications	
	außerhalb der Garantie / out of warranty	

Anerkannt / Accepted ☒	Nicht anerkannt / Not Accepted ☐
--	---

Austausch / Replacement

VB-Nr. / VB No

Seriennummer / Serial No.:

Lieferdatum / Delivery Date

Bearbeitet / Processed

04.02.22

Rücksendung / Return

Lieferdatum / Delivery Date

Geprüft / Inspected:

BMM

PRÜFBERICHT - TEST REPORT Sauerstoffsensoren MED / Oxygen Sensors Medical Use

Vorgangsnummer/ Process Number	47860588	Benachrichtigungsnummer/ Notification Number	300247850
-----------------------------------	----------	---	-----------



Kunde / Customer <i>Via med</i>	Kunden - Auftragsnummer / Customer Order No. RTS1100835
Auftragsdatum / Date of Order 05.01.2022	
Lief.an code/ Delivery Code	Aufgenommen von / Received by <i>KG</i>
Beanstandung / Objection <i>instabil</i>	

Artikelnummer / Part No.	Herstell. Datum / Date of Manufact.	Code + Seriennummer / Code + Serial No.	Garantie bis / End of Warranty Date	Beschreibung / Description
E1002780	06/21	108099	08/22	Oxygen Sensor OOM112 VIA

Prüfkriterien unter Standardbedingungen / Test Parameters under Standard Conditions

- | | | |
|---|---|------------------|
| 1 | Ausgangssignal an Luft / Output Signal in Atmosphere | [20,95 % vol.O2] |
| 2 | Linearitätsfehler, Meßfehler / Linearity Error, Error of Measurment | [100 % vol.O2] |
| 3 | Nullsignal / Zero Signal | [100 % vol.N2] |

Messprotokoll / Test Report

Prüfkriterien / Test Parameters	Spezifiziert / Specified	Gemessen / Measured	
		Signal1	Signal1
1 [mV]	25-50	33,7	
2 [%]	4400	7400	
3 [µV]	43	43	

Prüfbericht / Checklist

Fehler / Defect:	O2-S001..Luftwert außerh.Toleranz / Signal Out Of Tolerance	
	O2-S002..Signal Instabil / Signal Not Stable	
	O2-S003..Offset Außerh. Toleranz / Offset Out Of Tolerance	
	O2-S004..Signal 0mV / Signal 0mV	
	O2-S005..Gehäuse Defekt / Housing Defective	
	O2-S006..Elektrolytaustritt / Electrolyte Leakage	
	O2-S007..Sonstiges / Others	
Bemerkungen / Remark	kein Fehler festgestellt, Sensor innerhalb Spezifikation / No failure found, Sensor within specifications	
	außerhalb der Garantie / out of warranty	

Anerkannt / Accepted ☒	Nicht anerkannt / Not Accepted ☐
--	---

Austausch / Replacement	Rücksendung / Return
VB-Nr. / VB No	Lieferdatum / Delivery Date
Seriennummer / Serial No.:	Geprüft / Inspected: <i>B/M</i>
Lieferdatum / Delivery Date	
Bearbeitet / Processed	