

Service Repair Sheet SRS68080

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Alex Ferguson

John Radcliffe Hospital

Clinical Engineering

01865572163

01865 741166

00004040

Alex.Ferguson@ouh.nhs.uk

22/Feb/2022

Robert Connor

Viamed

Quote

Notes 18/Feb/2022 Sophie Lines

18/Feb/2022 Sophie Lines

Phoned up - has a faulty DB3 - had new batteries and leads, green LED lights up but nothing else. Gave him fixed price repair cost and SRS number - serial number is H0000254

22/Feb/2022 Robert Connor

1 x DB3 s/n H0000254, with patient lead and battery.

Ready For quote

C.Green 25.2.22

Repair Complete Signed

C.Green 21.3.22

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34743	Microstim DB3	2510000	H0000254	N

2540010 x 1 @ £45

S/N, SRS, SRN

UPS x 1 @ £10.

Service Repair Sheet 68080

Contact Name	Alex Ferguson
Company/ Hospital Name	John Radcliffe Hospital
Department	Clinical Engineering
Direct Phone	01865572163
General Phone	01865 741166
Opera Account	00004040
Email	Alex.Ferguson@ouh.nhs.uk
Order Number	RTHN400128570
Date Received	22/Feb/2022
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type
SRN34743	2510000	H0000254	Microstim DB3

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1	2540010
Qty: 1	2520000

The Microstim DB3 has been repaired and the patient lead has been replaced.
