

Certificate of Calibration

Certificate number:

Model Number:

Part Number:

Serial number:

This certificate reports recorded values for the instrument specified above.

The output values were measured in accordance with procedure:
VM3COP44.03.

Equipment used: Time Electronics LTD Microcal 1030 – Calibrated
Voltage source (CE076) - calibration traceable to
UKAS calibration standards.

Microcal output	MySign O expected reading ($\leq 2\%$ relative)	MySign O actual reading	Pass/Fail
50.0 mV	100% (98.0% - 102.0%)		
45.0 mV	90% (88.2% - 91.8%)		
40.0 mV	80% (78.4% - 81.6%)		
35.0 mV	70% (68.6% - 71.4%)		
30.0 mV	60% (58.8% - 61.2%)		
25.0 mV	50% (49.0% - 51.0%)		
20.0 mV	40% (39.2% - 40.8%)		
15.0 mV	30% (29.4% - 30.6%)		
10.5 mV	20.9% (20.5% - 21.3%)		
9.0 mV	18% (17.6% - 18.4%)		

Name:

Date:

Signed: _____