

Service Repair Sheet SRS68110

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Joanna Watkins

Eden Medical (UK) Ltd

0131 440 6500

0131 440 6500

00012178

joanna.watkins@edenmedical.co.uk

DPO-00865

22/Mar/2022

Robert Connor

Viamed

For Calibration

Notes 17/Mar/2022 Sarah Walton

17/Mar/2022 Sarah Walton

Calibration of theri V1000 unit S/N: PR02643A11

Advised them of Â£35.00 + Â£10.00 for calibration only and Â£45.00 + Â£10.00 for service and calibration.

22/Mar/2022 Robert Connor

1 x V1000 s/n PR02643A11, with 4 x AA battery in hard plastic carry case.

Purchase order supplied for 55 GBP, order no DPO-00865

Ready For quote

Repair Complete Signed

CGreen 23.3.22

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34818	Foetal Simulator	1410000	PR02643A11	N

1480000 x 1 @ £45  
S/N, SRS, SRN

1430309 x 1 £0 FOC  
SRS, SRN

UPS x 1 £10.

# PURCHASE ORDER

**Purchase Order To:**

Viamed Limited  
15 Station Road  
Cross Hills  
Keighley  
WEST YORKSHIRE  
BD20 7DT  
United Kingdom

**For Delivery To:**

Eden Medical (UK) Limited  
7 Dryden Place  
Bilston Glen Industrial Estate  
Loanhead  
Midlothian  
EH20 9HP  
United Kingdom

**Date**

21/03/2022

**Purchase Order No.**

DPO-00865

**VAT Registration No.**

GB 316 5811 10

Product Code	Order Qty	Product Description	Unit Cost	Unit Qty	Total
CUSTEQT	1	Customer Equipment for Service/Repair <i>Foetal Heart Rate Simulator V1000 [SN PR02643A11]</i>	45.00	EACH	45.00

**ADDITIONAL COSTS**

Description	Reference	Amount	Total
Standard carriage charge		10.00	10.00

Please note that all prices are shown in **GBP** and exclude VAT (if applicable).

**Total** 55.00

**NOTES:**

As per returns number SRS68110

**Important information:**

We reserve the right to reject any invoice or application for payment if any the details above have been altered without our agreement.  
Please note our order number must be quoted on all correspondence and a delivery note must accompany each consignment.

Page 1 of 1

7 Dryden Place, Bilston Glen Industrial Estate, Loanhead, Midlothian EH20 9HP  
Tel: 0131 440 6500 | Fax: 0131 202 0650 | Email: [info@edenmedical.co.uk](mailto:info@edenmedical.co.uk) | [www.edenmedical.co.uk](http://www.edenmedical.co.uk)  
Eden Medical (UK) Limited | Registered in Scotland No. SC345525 | Registered Office as above.



Service Repair Sheet 68110

Contact Name	Joanna Watkins
Company/ Hospital Name	Eden Medical (UK) Ltd
Direct Phone	0131 440 6500
General Phone	0131 440 6500
Opera Account	00012178
Email	joanna.watkins@edenmedical.co.uk
Order Number	DPO-00865
Date Received	22/Mar/2022
Booked in By	Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN34818	1410000	PR02643A11	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked

---