

AIC on stop
cg 24-3-22

Service Repair Sheet SRS68098

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Karl Green

Company/ Hospital Name

Alexandra Hospital

Department

Position

Direct Phone

01527503030

General Phone

01527 503030

Opera Account

00004295

Email

karl.green1@nhs.net

Order Number

Date Received

10/Mar/2022

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Quote

Notes 07/Mar/2022 Zoey Teal

07/Mar/2022 Zoey Teal

Sending in a V1000 for service and calibration. Serial number PR02913A10

10/Mar/2022 Robert Connor

Received 1 x V1000 s/n PR02913A10, with grey silicone cover and blue fabric carry case.

Ready For quote

Cgreen 11-3-22.

Repair Complete Signed

Cgreen 24-3-22.

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34788	Foetal Simulator	1410000	PR02913A10	N

1480000 x 1 @ £45
S/N, SRS, SRN

1430309 x 1 FOC.
SRS, SRN

UPS x 1 @ £10.



Supplier:

VIAMED LTD
15 STATION ROAD
CROSS HILLS

KEIGHLEY, WEST YORKSHIRE BD20 7DT

01535634542

GLN:

Buyer MICHAEL RWP PAUL

Telephone

Email m.paul2@nhs.net

RWP 182848 TECHNICAL SERV DEPT

Deliver to:

KIDDERMINSTER HOSPITAL
RECEIPTS AND DISTRIBUTION
FRANCHISE STREET
KIDDERMINSTER, DY11 6RJ

Invoice to:

WORCESTERSHIRE ACUTE HOSPITAL
RWP PAYABLES 6485
PHOENIX HOUSE TOPCLIFFE LANE
WAKEFIELD, WF3 1WE0303 123 1177
GLN:

Order Number 305469528

Date 14-MAR-22

PLEASE CHECK THIS P.O. IF ANY OF THE DETAILS CONCERNING THE ITEMS LISTED ARE BELIEVED TO BE INCORRECT, I.E. SUPPLIER CODE, ITEM DESCRIPTION, PRICE OR DELIVERY CHARGE.

PLEASE EMAIL FULL DETAILS OF THE AMENDMENTS TO:

WAH-TR.PURCHASING@NHS.NET OR ALTERNATIVELY FAX AN AMENDED COPY OF THE P.O. TO 01527 502822 AND,

IF REQUIRED, THE ORDER WILL BE RESUBMITTED TO YOU.

Quantity Required	U.O.M	Supplier Part Number:	Description	Delivery Date	Unit Price (Inc Discount)	Line Value GBP
		calibration	Calibration to Foetal Heart simulator test equipmentQuote number QVM135923	21-MAR-22		45.00
		Shipping	Return shipping for Foetal Heart simulator test equipmentQuote number QVM135923	21-MAR-22		10.00

Total Value of Order (Exc VAT)

55.00

Instructions to Supplier: This order is subject to the standard NHS Terms and Conditions of contract. For a copy of these please contact the Buyer for this order. Any price alterations must be agreed with the buyer prior to order execution. The above order number must be quoted on all invoices, acknowledgements, delivery notes and other correspondence. A delivery note must accompany each consignment of goods. The order must not be passed to any third party for supply. Any invoices not complying with these instructions will be returned unpaid to the supplier.

Service Repair Sheet 68098

Contact Name	Karl Green
Company/ Hospital Name	Alexandra Hospital
Direct Phone	01527503030
General Phone	01527 503030
Opera Account	00004295
Email	karl.green1@nhs.net
Order Number	305469528
Date Received	10/Mar/2022
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type
SRN34788	1410000	PR02913A10	Foetal Simulator

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1	1430309
• Qty: 1	1480000

V1000 has had transducer interface cushion replaced and calibration has been checked
