

Service Repair Sheet SRS68118

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Chandra Suresh

Company/ Hospital Name

Royal Free Hospital

Department

Medical Electronics Department

Position

Bio-Medical Engineer

Direct Phone

0207 8302417

General Phone

020 3322 1935

Opera Account

00003220

Email

Chandra.Sureshkumar@nhs.net

Order Number

Date Received

28/Mar/2022

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Quote

Notes 28/Mar/2022 Robert Connor

28/Mar/2022 Robert Connor

Received 1 x VM-2500-S s/n 81040152, with silicone cover and li-poly battery. Unit screen is smashed.

Customer requests quote for screen repair, calibration, and replacement probe.

Ready For quote

CGreen 4-4-22

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34927	Capnograph	4410520	81040152	N

4440000 x 1/2 hr £90 = £45
S/N, SRS, SRN

4430020 x 1 @ £9.
SRS, SRN

4480520 x 1 @ £150
SRS, SRN

UPS x 1 @ £10.



Request for Repair Estimate

Company Name:	Viamed Ltd	RFH Contact:	Chandra Suresh
Attention:	Zoey Teal, Tech Support	RFH Contact E-mail:	Chandra.sureshkumar@nsh.net
Company Reference		Phone Number:	020 7830 2417
Date:	24 March 2022		

The following equipment has been sent for an estimate for repair. Please contact the above listed member of Medical Electronics with your quotation. Do not proceed with the repair until you have received an official order from the trust. This will originate from the Partners Procurement Service. Any invoice raised without a purchase order may not be approved.

RFH Number	Job Number	Model Number	Serial Number	Reported Problem
98288	J344376	VM-2500-s	81040152	Broken screen needs repair. missing sensor probe needs replacement. Also needs calibration.

Comments:

Delivery Address:
(**DO NOT** send to General Stores)

Medical Electronics Department,
Royal Free Hospital,
Pond Street,
London,
NW3 2QG

Invoice Address:
Please see official order

Decontamination: The equipment has been decontaminated by the users prior to shipping. If you require a copy of this certificate please contact us quoting the job number.

Information Governance: If Personal Identifiable Data is found on this device please ensure you follow Schedule 3 of *NHS Terms and Conditions for the Supply of Goods (Purchase Order Version)*. Please contact us if you require further guidance.

NHS Terms and Conditions apply

NHS terms and conditions are available from website www.dh.gov.uk

Catherine Spence Required Reading 1 Calls 42 Issues 28 Unread 28 ISO Tasks |
 H/W: CS DL GB HL MG PC Future Stock



Select Currency

Pounds | Euros | Usd

Repair SRS68118 Values in Pounds @ £1: £1.00

SRN34927	Capnograph	4410520			
	Parts Required	Reference	Quantity	Unit Cost	Total Line
		4480520	1	£150.00	£150.00
		4430020	1	£9.90	£9.90
		4440000	0.5	£90.00	£45.00
				Parts Cost	£204.90

CHECK PRICES BEFORE SENDING QUOTE

Catherine Spence Required Reading 1 Calls 42 Issues 28 Unread 28 ISO Tasks |
 H/W: CS DL GB HL MG PC Future Active Users Stock Search



Select Currency

Pounds| Euros| Usd

Repair SRS68118 Values in Pounds @ £1: £1.00

SRN34927	Capnograph	4410520			
	Parts Required	Reference	Quantity	Unit Cost	Total Line
		4480520	1	£150.00	£150.00
		4430020	1	£9.90	£9.90
		4440000	0.5	£90.00	£45.00
				Parts Cost	£204.90

CHECK PRICES BEFORE SENDING QUOTE

Quote Sent