

Service Repair Sheet SRS68123

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Ashley Davies

Thames Medical Ltd

Office Manager

01903522911

00011982

ashley@thamesmedical.com

PO-01737

31/Mar/2022

Robert Connor

Viamed

Quote

Notes 29/Mar/2022 Kate Griffiths

29/Mar/2022 Kate Griffiths

Customer to send "in potential screen replacement on a VM2160, serial number " 70600146".

31/Mar/2022 Robert Connor

1 x VM2160 s/n 70600146, with clip sensor EDE00814, and 3 x AA battery. Battery contacts rusty.

Ready For quote

CGreen 4-4-22.

Repair Complete Signed

CGreen 21-4-22

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34933	Pulse Oximeter	0012175	70600146	N
SRN34934	Pulse Oximetry Ear Clip Sensor	0015000	EDE00814	N

0039031 x 1 @ £89.
~~SRN~~, SRS, SRN.

0039030 x 1 @ £0.
SRS, SRN

0039002 x 1 @ £4.50
SRS, SRN.

0040001 x 1 @ £0. - SRN 34934
S/N, SRN, SRS.
UPS x 1 @ £10



PURCHASE ORDER

order # PO-01737

date 29/03/2022

Thames Medical Ltd
The Garden
Rushy, West Sussex BN10 3BW
www.thamesmedical.com
email: info@thamesmedical.com
tel: 01903 522922
fax: 01903 522922
Experts in patient monitoring and anaesthesia

vendor address: Vamed Ltd
15 Station Road
Cross Hills
Keighley, West Yorkshire
BD20 7DT

contact: Kate
phone: 01535 634542
fax: 01535 635582

carrier

PS

payment terms

Net 30

ITEM	DESCRIPTION	VENDOR PRODUCT CODE	QUANTITY	UNIT PRICE	SPECIAL TAX RATE	SUB-TOTAL
TML-V-VM2160 RL2WR	VM2160 - Warranty Repair - Service Level 2 SN-70600146	Warranty Repair of VM2160 Pulse Oximeter	1	£0.00	Taxable	£0.00

remarks	SRS68123	SUB-TOTAL	£0.00
	Please quote before any work takes place.	VAT	£0.00
		TOTAL	£0.00

not under warranty
cg 4.4.22

Service Repair Sheet 68123

Contact Name	Ashley Davies
Company/ Hospital Name	Thames Medical Ltd
Position	Office Manager
Direct Phone	01903522911
Opera Account	00011982
Email	ashley@thamesmedical.com
Order Number	PO-01737
Date Received	31/Mar/2022
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN34933	0012175	70600146	Pulse Oximeter	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	0039002
Qty: 1	0039031
Qty: 1	0039030

The VM2160 Vet pulse oximeter has been repaired. The TFT upgrade kit has been installed. The battery contacts have been replaced.

Repair	Ref	S/N	Equipment Type	
SRN34934	0015000	EDE00814	Pulse Oximetry Ear Clip Sensor	No Fault Found Time :0 Hour(s)
