

Service Repair Sheet SRS68123  
 Contact Name  
 Company/ Hospital Name  
 Department  
 Position  
 Direct Phone  
 General Phone  
 Opera Account  
 Email  
 Order Number  
 Date Received  
 Booked in By  
 Main Company  
 Type Return

Date Promised to Return: 01 Jan 1970 -3600  
 Ashley Davies  
 Thames Medical Ltd  
 Office Manager  
 01903522911  
 00011982  
 ashley@thamesmedical.com  
 PO-01737  
 31/Mar/2022  
 Robert Connor  
 Viamed  
 Quote

Notes 29/Mar/2022 Kate Griffiths

29/Mar/2022 Kate Griffiths

Customer to send "in potential screen replacement on a VM2160, serial number " 70600146".

31/Mar/2022 Robert Connor

1 x VM2160 s/n 70600146, with clip sensor EDE00814, and 3 x AA battery. Battery contacts rusty.

Ready For quote

CGreen 4-4-22.

Repair Complete Signed

CGreen 21-4-22

| SRN      | Equipment                      | Stock Ref | Serial Number | Warranty |
|----------|--------------------------------|-----------|---------------|----------|
| SRN34933 | Pulse Oximeter                 | 0012175   | 70600146      | N        |
| SRN34934 | Pulse Oximetry Ear Clip Sensor | 0015000   | EDE00814      | N        |

0042160 x 1 FOC  
 S/N, SRS, SRN

-> Added by Cg, n  
 off by mistake  
 cg 21-4-22

0039031 x 1 @ £89.  
 S/N, SRS, SRN.

0039030 x 1 @ £0.  
 SRS, SRN

0039002 x 1 @ £4.50  
 SRS, SRN.

0040001 x 1 @ £0. - SRN 34934  
 S/N, SRN, SRS.  
 UPS x 1 @ £10