

Invoice Correction 5259728883

ORIGINAL

Invoice Correction Date: 11.04.2022

Honeywell Healthcare Solutions GmbH

Purchase Order No: RTS3100824

Alter Holzhafen 18

23966 Wismar

Germany

Bill to: 1797471

Vandagraph Limited
15 Station Road
Cross Hills, West Yorkshire
BD20 7DT
KEIGHLEY
United Kingdom

Sold to: 1797448

Vandagraph Sensor Technologies Ltd.
15 Station Road
Cross Hills, West Yorkshire
KEIGHLEY
BD20 7DT
United Kingdom

Order Information:

Credit Memo Request: 63854977
Sales Office: 2170 DE SPS Wismar
Incoterms: FCA Origin
Payment Terms: Net 30 Days After Invoice Date

Ship to: 1797448

Vandagraph Sensor Technologies Ltd.
15 Station Road
Cross Hills, West Yorkshire
KEIGHLEY
BD20 7DT
United Kingdom

ITEM NO.	Material No.	QTY	UOM	UNIT PRICE	EXT. PRICE	VAT %
000010	E1001581 Oxygen Sensor OOD103-1V Customer Part Nbr: 8010007 credit note for SN: 114471; 114947	2	EA	44,44	-88,88	0%
				<i>Should be 1x €20.73 1x €20.00</i>		
000020	E1002226 Oxygen Sensor OOD103-APD16 Customer Part Nbr: 8010022 credit note for SN: 119642; 120818; 120962	3	EA	44,44	-133,32	0%
				<i>Should be 1x €20 2x €20.73 cg 21-6-22</i>		

Totals

Total Before VAT -222,20 EUR
VAT-Total 0,00 EUR
Total Amount -222,20 EUR

Tax Information

Tax Exempt Reason: Tax free export - Goods
Fulfilment Date: 11.04.2022

Duns #:

Bank Details: No Payment Due

Account Name:

Account No:

Bank Key:

Swift Code:

IBAN:

Payment method:

Registered No: Schwerin HRB 1838

VAT Reg No: DE137437213

Federal ID #:

Registered Address: Honeywell Healthcare Solutions GmbH, Alter Holzhafen 18, Wismar, 23966, Germany

Credit Analyst: Xenia Musat

Sales Rep: Stefan Dreyer

Customer Care Rep: Jessica Wagner + 49 3841 360 206

Honeywell Healthcare Solutions GmbH

WEEE Reg.No.: 57798680
Managing Director: Marcus Ostländer

Internet: www.envitec.com
email: info@envitec.com

Delivery exclusive to our General Terms (February 2016). We retain title to our goods and deliveries until all the payment from the sales contract have been received. Your order will be executed exclusively to our terms and conditions. These can be sent by separate request or viewed at www.envitec.com

PRÜFBERICHT - TEST REPORT

Sauerstoffsensoren/Oxygen Sensors OOA, OOI, OOD

ENVITEC
by Honeywell

Kunde/ Customer	Vauda graph	Retour/ RMA- No.	Kunde/ Customer	RTS 3100P24	Envitec	300245-390
Artikel/ Article	000103-11; 000103-A PD ₆	Anzahl/ Quantity	2; 3	Eing. Datum/ Inc. Date	01.11.21	
Bearbeitet/ processed	U. B. m	Prüfdatum/ Insp. Date	8.11.21	Geprüft/ Inspected	Bill	

Produkt/ Identification			Messwert/ Test Result				Bewertung/ Evaluation			
Nr.	Seriennummer/ Serial no.	Lieferlos/ Manufact. code	Luftwert/ Air signal @ 20,9% O ₂ [mV]	Nullsignal/ Offset @ 0% O ₂ [µV]	Linearität/ Linearity [%]	Bemerkung/ Remark	außerhalb Garantie/ out of warranty	kein Fehler festgestellt/ no failure found	Fehlercode/ Failure- code	Austausch / Replacement
1	114471	05121	11.3	27	i.o.	Drucktest			07	
2	114947	08121	11.1	28	i.o.			x		
3	119642	11120	10.7	11	i.o.			x		
4	120818	09121	11.8	9	i.o.			x		
5	120962	09121	11.7	10	i.o.			x		
6										
7										
8										
9										
10										
11										
12										
13										
14										
15										
16										
17										
18										
19										
20										

Fehler O2-S001: Luftwert außerhalb Toleranz/ Air signal out of tolerance
 Fehler O2-S003: Offset außerhalb Toleranz/ Offset out of tolerance
 Fehler O2-S005: Gehäuse defekt/ Housing defective
 Fehler O2-S007: Sonstiges (unter Angabe in Bemerkung) / Others (see remark)

Fehler O2-S002: Signal instabil / Signal not stable
 Fehler O2-S004: Signal 0mV / Signal 0mV
 Fehler O2-S006: Elektrolyt Austritt / Electrolyte leakage