

Service Repair Sheet SRS68138

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Ashley Davies

Company/ Hospital Name

Thames Medical Ltd

Department

Position

Direct Phone

01903522911

General Phone

01903 522911

Opera Account

00011982

Email

ashley@thamesmedical.com

Order Number

Date Received

21/Apr/2022

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Quote

Notes 20/Apr/2022 Zoey Teal

20/Apr/2022 Zoey Teal

Sending in a IRMA with serial number 259304. Note from customer - It is reporting a fault and also has a cracked connector, we just need a quote for potential repair with it being more likely the client will replace.

21/Apr/2022 Robert Connor

IRMA analyser s/n 259304, mini med connector broken.

Ready For quote

CG Neen 25.4.2022

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34981	Accessory	4420505	259304	N

~~44400010~~

4420507

~~4420505~~ x 1

SRS, SRN

£905.00

@ £807.50

cg 26.4.2022

UPS x 1 @ £10.