

Device Essential Requirements Checklist
(in compliance with the 93/42/EEC Medical Device Directive).

Product:		New product:	
Part number:		Existing Product:	
Description:		Introduced:	
Class:		Main Standard:	

No.	Essential Requirement.	A/NA.	Standard.	Report.
I	General Requirements			
1.	The devices must be designed and manufactured in such a way that, when used under the conditions and for the purposes intended, they will not compromise the clinical condition or the safety of patients, or the safety and health of users or, where applicable, other persons, provided that any risks which may be associated with their use constitute acceptable risks when weighed against the benefits to the patient and are compatible with a high level of protection of health and safety.	A	IEC 601/1	Manufactured to ISO 9000 quality standards. (E) Risk analysis. (YZ) Design File
2.	The solutions adopted by the manufacturer for the design and construction of the devices must conform to safety principles, taking account of the generally acknowledged state of the art.	A	BS EN ISO 14971:2001	Compatibility with OEM Product (E) Risk analysis. Known risks reduced for hazards identified
	- Eliminate or reduce risks as far as possible (inherently safe design and construction),	A		(E) Risk analysis
	- Where appropriate take adequate protection measures including alarms if necessary, in relation to risks that cannot be eliminated,	A		(E) Risk analysis
	- Inform users of the residual risks due to any shortcomings of the protection measures adopted	A	BS EN 980 EN1041	(E) Risk analysis (F) User manual
3.	The devices must achieve the performances intended by the manufacturer and be designed, manufactured and packaged in such a way that they are suitable for one or more of the functions referred to in Article 1 (2) (a), as specified by the manufacturer.	A		(YZ) Specification. (YZ) Validation. (YZ) Clinical trials. (YZ) Tests. (M) Packaging trials & validation.
4.	The characteristics and performances referred to in Sections 1, 2 and 3 must not be adversely affected to such a degree that the clinical conditions and safety of the patients and, where applicable, of other persons are compromised during the lifetime of the device as indicated by	A	BS EN ISO 14971:2001.	(E) Risk analysis Long term Clinical Trials

	the manufacturer, when the device is subjected to the stresses which can occur during normal conditions of use.			
5.	The devices must be designed, manufactured and packed in such a way that their characteristics and performances during their intended use will not be adversely affected during transport and storage taking account of the instructions and information provided by the manufacturer.	A		(M) Packaging trials & validation (T) Manufacturers data & Materials Specifications Specific Packaging Design
6.	Any undesirable side effect must constitute an acceptable risk when weighed against the performances intended.	N/A	BS EN ISO 14971:2001	No known side effects (E) Risk analysis
II	Requirements Regarding Construction & Design			
7.	Chemical, physical and biological properties			
7.1	The devices must be designed and manufactured in such a way as to guarantee the characteristics and performances referred to in Section I on the 'General requirements'. Particular attention must be paid to:	A		Medical grade materials used Manufacturers data (T) Material specifications
	-The choice of materials used, particularly as regards toxicity and, where appropriate, flammability,	A		Non-inflammable (YZ) Design Reviews (T) Material specifications
	- The compatibility between the materials used and biological tissues, cells and body fluids, taking account of the intended purpose of the device.	A		Medical grade materials used (T) Material specifications
7.2	The devices must be designed, manufactured and packed in such a way as to minimize the risk posed by contaminants and residues to the persons involved in the transport, storage and use of the devices and to the patients, taking account of the intended purpose of the product. Particular attention must be paid to the tissues exposed and to the duration and frequency of exposure.	A		Manufactured to ISO 9000 quality standards. (E) Risk analysis (M) Packaging trials & validation
7.3	The devices must be designed and manufactured in such a way that they can be used safely with the materials, substances and gases with which they enter into contact during their normal use or during routine procedures; if the devices are intended to administer medicinal products they must be designed and manufactured in such a way as to be compatible with the medicinal products concerned according to the provisions and restrictions	A		(YZ) Clinical Trials (YZ) Test Reports (T) Material specifications – materials identical to OEM

	governing these products and that their performance is maintained in accordance with the intended use.			
7.4	Where a device incorporates, as an integral part, a substance which, if used separately, may be considered to be a medicinal product as defined in Article 1 of Directive 65/65/EEC and which is liable to act upon the body with action ancillary to that of the device, the safety, quality and usefulness of the substance must be verified, taking account of the intended purpose of the device, by analogy with the appropriate methods specified in Directive 75/318/EEC.	N/A		No Medicinal Products
7.5	The devices must be designed and manufactured in such a way as to reduce to a minimum the risks posed by substances leaking from the device.			Nothing to Leak
7.6	Devices must be designed and manufactured in such a way as to reduce, as much as possible, risks posed by the unintentional ingress of substances into the device taking into account the device and the nature of the environment in which it is intended to be used.	A	BS EN ISO 14971:2001	(E) Risk analysis Manufacturers Data (YZ) Design Reviews (T) Material specifications
8	Infection and microbial contamination			
8.1	The devices and manufacturing processes must be designed in such a way as to eliminate or reduce as far as possible the risk of infection to the patient, user and third parties. The design must allow easy handling and, where necessary, minimize contamination of the device by the patient or vice versa during use.	A	ISO900:2000	Manufacturing procedures. Cleaning Instructions Manufacturers Data (T) Material specifications
8.2	Tissues of animal origin must originate from animals that have been subjected to veterinary controls and surveillance adapted to the intended use of the tissues. Notified bodies shall retain information on the geographical origin of the animals. Processing, preservation, testing and handling of tissues, cells and substances of animal origin must be carried out so as to provide optimal security. In particular safety with regard to viruses and other transferable agents must be addressed by implementation of validated methods of elimination or viral inactivation in the course of the manufacturing process.	N/A		No animal origin components
8.3	Devices delivered in a sterile state must be designed, manufactured and packed in a non-reusable pack and/or according to appropriate procedures to ensure that they are sterile when placed on the market and remain sterile, under the storage and transport conditions laid down,	N/A		Non Sterile

	until the protective packaging is damaged or opened.			
8.4	Devices delivered in a sterile state must have been manufactured and sterilized by an appropriate, validated method.	N/A		Not supplied Sterile
8.5	Devices intended to be sterilized must be manufactured in appropriately controlled (e. g. environmental) conditions.	N/A		Cannot be Sterilised
8.6	Packaging systems for non-sterile devices must keep the product without deterioration at the level of cleanliness stipulated and, if the devices are to be sterilized prior to use, minimize the risk of microbial contamination; the packaging system must be suitable taking account of the method of sterilization indicated by the manufacturer.	A		(M) Packaging
8.7	The packaging and/or label of the device must distinguish between identical or similar products sold in both sterile and non-sterile condition.	N/A		Non Sterile
9	Construction and environmental properties			
9.1	If the device is intended for use in combination with other devices or equipment, the whole combination, including the connection system must be safe and must not impair the specified performances of the devices. Any restrictions on use must be indicated on the label or in the instructions for use.	A	BS EN 980 ISO 9001:2000 BS EN ISO 14971:2001	Device fully tested for compatibility (F) User manual (E) Risk analysis
9.2	Devices must be designed and manufactured in such a way as to remove or minimize as far as is possible: - The risk of injury, in connection with their physical features, including the volume/pressure ratio, dimensional and where appropriate ergonomic features,	N/A		
	- Risks connected with reasonably foreseeable environmental conditions, such as magnetic fields, external electrical influences, electrostatic discharge, pressure, temperature or variations in pressure and acceleration,	A		EMC proofed and tested
	- The risks of reciprocal interference with other devices normally used in the investigations or for the treatment given,	A		EMC proofed and tested
	- risks arising where maintenance or calibration are not possible (as with implants), from ageing of materials used or loss of accuracy of any measuring or control mechanism.	N/A		No calibration required Ageing at present under review Not conclusive
9.3	Devices must be designed and manufactured in such a way as to minimize the risks of fire or	N/A		Non inflammable

	explosion during normal use and in single fault condition. Particular attention must be paid to devices whose intended use includes exposure to flammable substances or to substances, which could cause combustion.			
10	Devices with a measuring function			
10.1	Devices with a measuring function must be designed and manufactured in such a way as to provide sufficient accuracy and stability within appropriate limits of accuracy and taking account of the intended purpose of the device. The manufacturer must indicate the limits of accuracy.	A	ISO 9001:2000	(YZ) Specification (YZ) Design history (YZ) Design reviews (YZ) Clinical trials (YZ) Test reports
10.2	The measurement, monitoring and display scale must be designed in line with ergonomic principles, taking account of the intended purpose of the device.	N/A		Accessory
10.3	The measurements made by devices with a measuring function must be expressed in legal units conforming to the provisions of Council Directive 80/181/EEC (1).	N/A		Accessory
11	Protection against radiation			
11.1	General			
11.1	Devices shall be designed and manufactured in such a way that exposure of patients; users and other persons to radiation shall be reduced as far as possible compatible with the intended purpose, whilst not restricting the application of appropriate specified levels for therapeutic and diagnostic purposes.	N/A		No Ionizing radiation
11.2	Intended radiation			
11.2.1	Where devices are designed to emit hazardous levels of radiation necessary for a specific medical purpose the benefit of which is considered to outweigh the risks inherent in the emission, it must be possible for the user to control the emissions. Such devices shall be designed and manufactured to ensure reproducibility and tolerance of relevant parameters.	N/A		No Ionizing radiation
11.2.2	Where devices are intended to emit potentially hazardous, visible and/or invisible radiation, they must be fitted, where practicable, with visual displays and/or audible warnings of such emissions.	N/A		No radiation
11.3	Unintended radiation			
11.3.1	Devices shall be designed and manufactured in such a way that exposure of patients, users and other persons to the emission of unintended, stray or scattered radiation is reduced as far as	N/A		No radiation

	possible.			
11.4	Instructions.			
11.4.1	The operating instructions for devices emitting radiation must give detailed information as to the nature of the emitted radiation, means of protecting the patient and the user and on ways of avoiding misuse and of eliminating the risks inherent in installation.	N/A		No radiation
11.5	Ionizing radiation			
11.5.1	Devices intended to emit ionizing radiation must be designed and manufactured in such a way as to ensure that, where practicable, the quantity, geometry and quality of radiation emitted can be varied and controlled taking into account the intended use.	N/A		No Ionizing radiation
11.5.2	Devices emitting ionizing radiation intended for diagnostic radiology shall be designed and manufactured in such a way as to achieve appropriate image and/or output quality for the intended medical purpose whilst minimizing radiation exposure of the patient and user.	N/A		No Ionizing radiation
11.5.3	Devices emitting ionizing radiation, intended for therapeutic radiology shall be designed and manufactured in such a way as to enable reliable monitoring and control of the delivered dose, the beam type and energy and where appropriate the quality of radiation.	N/A		No Ionizing radiation
12	Requirements for medical devices connected to or equipped with an energy source			
12.1	Devices incorporating electronic programmable systems must be designed to ensure the repeatability, reliability and performance of these systems according to the intended use. In the event of a single fault condition (in the system) appropriate means should be adopted to eliminate or reduce as far as possible consequent risks.	N/A		No software
12.2	Devices where the safety of the patients depends on an internal power supply must be equipped with a means of determining the state of the power supply.	N/A		Accessory
12.3	Devices where the safety of the patients depends on an external power supply must include an alarm system to signal any power failure.	N/A		Accessory
12.4	Devices intended to monitor one or more clinical parameters of a patient must be equipped with appropriate alarm systems to alert the user of situations which could lead to death or severe deterioration of the patient's	N/A		Accessory

	state of health.			
12.5	Devices must be designed and manufactured in such a way as to minimize the risks of creating electromagnetic fields which could impair the operation of other devices or equipment in the usual environment.	A	BS 60601-1-2	Shielded
12.6	Protection against electrical risks			
12.6.1	Devices must be designed and manufactured in such a way as to avoid, as far as possible, the risk of accidental electric shocks during normal use and in single fault condition, provided the devices are installed correctly.	A		LED's insulated Protection in Oximeter
12.7	Protection against mechanical and thermal risks			
12.7.1	Devices must be designed and manufactured in such a way as to protect the patient and user against mechanical risks connected with, for example, resistance, stability and moving parts.	A		Designed to prevent excessive movement
12.7.2	Devices must be designed and manufactured in such a way as to reduce to the lowest possible level the risks arising from vibration generated by the devices, taking account of technical progress and of the means available for limiting vibrations, particularly at source, unless the vibrations are part of the specified performance.	N/A		No vibration
12.7.3	Devices must be designed and manufactured in such a way as to reduce to the lowest possible level the risks arising from the noise emitted taking account of technical progress and of the means available to reduce noise, particularly at source, unless the noise emitted is part of the specified performance.	N/A		No noise emitted
12.7.4	Terminals and connectors to the electricity, gas or hydraulic and pneumatic energy supplies which the user has to handle must be designed and constructed in such a way as to minimize all possible risks.	N/A		No connections involved
12.7.5	Accessible parts of the devices (excluding the parts or areas intended to supply or reach given temperatures) and their surroundings must not attain potentially dangerous temperatures under normal use.	N/A		Low temperature only
12.8	Protection against the risks posed to the patient by energy supplies or substances			
12.8.1	Devices for supplying the patient with energy or substances must be designed and constructed in such a way that the flow-rate can be set and maintained accurately enough to guarantee the safety of the patient and of the user.	N/A		No flow rates involved
12.8.2	Devices must be fitted with the means of preventing and/or indicating any inadequacies	N/A		No flow rates involved

	in the flow-rate, which could pose a danger. Devices must incorporate suitable means to prevent, as far as possible, the accidental release of dangerous levels of energy from an energy and/or substance source.			
12.9	The function of the controls and indicators must be clearly specified on the devices. Where a device bears instructions required for its operation or indicates operating or adjustment parameters by means of a visual system, such information must be understandable to the user and, as appropriate, the patient.	N/A		Accessory
13	Information supplied by the manufacturer			
13.1	Each device must be accompanied by the information needed to use it safely and to identify the manufacturer, taking account of the training and knowledge of the potential users. This information comprises the details on the label and the data in the instructions for use. As far as practicable and appropriate, the information needed to use the device safely must be set out on the device itself and/or on the packaging for each unit or, where appropriate, on the sales packaging. If individual packaging of each unit is not practicable, the information must be set out in the leaflet supplied with one or more devices. Instructions for use must be included in the packaging for every device. By way of exception, no such instructions for use are needed for devices in Class I or II(a) if they can be used safely without any such instructions.	A	BS EN 980. EN1041	Information supplied by the manufacturer with medical devices (F) User manual,
13.2	Where appropriate, this information should take the form of symbols. Any symbol or identification color used must conform to the harmonized standards. In areas for which no standards exist, the symbols and colors must be described in the documentation supplied with the device.	A	BS EN 980. EN1041	Graphic symbols for use in the labeling of medical devices (F) Labels
13.3	The label must bear the following particulars	A	BS EN 980 EN1041	Graphic symbols for use in the labeling of medical devices All controls are marked (F) labels
(a)	The name or trade name and address of the manufacturer. For devices imported into the Community, in view of their distribution in the Community, the label, or the outer packaging, or instructions for use, shall contain in addition	A	BS EN 980 EN1041	(F) Label (F) User manual

	the name and address of either the person responsible referred to in Article 14 (2) or of the authorized representative of the manufacturer established within the Community or of the importer established within the Community, as appropriate;			
(b)	The details strictly necessary for the user to identify the device and the contents of the packaging;	A	BS EN 980 EN1041	(F) Label (F) User manual
(c)	Where appropriate, the word 'STERILE';	N/A		Not Sterile
(d)	Where appropriate, the batch code, preceded by the word 'LOT', or the serial number;	A	BS EN 980 EN1041	(F) Label (F) User manual (F) Serial number label
(e)	Where appropriate, an indication of the date by which the device should be used, in safety, expressed as the year and month;	N/A		Not required – No end of life set
(f)	Where appropriate, an indication that the device is for single use;	N/A		Re-usable
(g)	If the device is custom-made, the words 'custom-made device';	N/A		
(h)	If the device is intended for clinical investigations, the words 'exclusively for clinical investigation';	N/A		For clinical use
(i)	Any special storage and/or handling conditions;	N/A		None required
(j)	Any special operating instructions;	A	BS EN 980	(F) User manual
(k)	Any warnings and/or precautions to take;	A	BS EN 980	(F) User manual
(l)	Year of manufacture for active devices other than those covered by (c). This indication may be included in the batch or serial number;	N/A		
(m)	Where applicable, method of sterilization.	N/A		Not to be Sterilised
13.4	If the intended purpose of the device is not obvious to the user, the manufacturer must clearly state it on the label and in the instructions for use.	A	BS EN 980 EN1041	Intended purpose is obvious (F) User manual
13.5	Wherever reasonable and practicable, the devices and detachable components must be identified, where appropriate in terms of batches, to allow all appropriate action to detect any potential risk posed by the devices and detachable components	N/A		
13.6	Where appropriate, the instructions for use must contain the following particulars:			
(a)	The details referred to in Section 13.3, with the exception of (d)&(c);	A	BS EN 980 EN1041	Full Instruction leaflet supplied (F) User manual
(b)	The performances referred to in Section 3 and any undesirable side-effects;	N/A		No side effects
(c)	If the device must be installed with or connected to other medical devices or	A	BS EN 980 EN1041	Full Instruction leaflet supplied

	equipment in order to operate as required for its intended purpose, sufficient details of its characteristics to identify the correct devices or equipment to use in order to obtain a safe combination;			(F) User manual
(d)	All the information needed to verify whether the device is properly installed and can operate correctly and safely, plus details of the nature and frequency of the maintenance and calibration needed to ensure that the devices operate properly and safely at all times;	N/A		
(e)	Where appropriate information to avoid certain risks in connection with implantation of the device;	N/A		Not implantable
(f)	Information regarding the risks of reciprocal interference posed by the presence of the device during specific investigations or treatment;	N/A		
(g)	The necessary instructions in the event of damage of the sterile packaging and, where appropriate, details of appropriate methods of re-sterilization;	N/A		Not sterile
(h)	If the device is reusable, information on the appropriate processes to allow reuse, including cleaning, disinfection, packaging and, where appropriate, the method of sterilization of the device to be re-sterilized, and any restriction on the number of reuses.	A	BS EN 980 EN1041	Full Instruction leaflet supplied (F) User manual
	Where devices are supplied with the intention that they be sterilized before use, the instructions for cleaning and sterilization must be such that, if correctly followed, the device will still comply with the requirements in Section I;	N/A		Not Sterile
(i)	Details of any further treatment or handling needed before the device can be used (for example, sterilization, final assembly, etc.);	N/A		No sterilisation
(j)	In the case of devices emitting radiation for medical purposes, details of the nature, type, intensity and distribution of this radiation. The instructions for use must also include details allowing the medical staff to brief the patient on any contra-indications and any precautions to be taken. These details should cover in particular:	N/A		No radiation
(k)	Precautions to be taken in the event of changes in the performance of the device;	N/A		
(l)	Precautions to be taken as regards exposure, in reasonably foreseeable environmental conditions, to magnetic fields, external electrical influence, electrostatic discharge,	A	BS EN 980 EN1041	MRI Advice in Leaflet Full Instruction leaflet supplied

	pressure or variations in pressure, acceleration, thermal ignition sources, etc.;			(F) User manual
(m)	Adequate information regarding the medicinal product or products which the device in question is designed to administer, including any limitations in the choice of substances to be delivered;	N/A		Does not administer substances
(n)	Precautions to be taken against any special, unusual risks related to the disposal of the device;	N/A		No risks in disposable of the device
(o)	Medicinal substances incorporated into the device as an integral part in accordance with section 7.4;	N/A		No medicinal products used
(p)	Degree of accuracy claimed for devices with a measuring function.	A		(YZ) Specification (F) User Manual
14.	Where conformity with the essential requirements must be based on clinical data, as in Section I (6), such data must be established in accordance with Annex X.	A		Compatibility tests only required