

Service Repair Sheet SRS68087

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Sean Bunn

Company/ Hospital Name

Luton & Dunstable Hospital

Department

Position

Direct Phone

01582718369

General Phone

01582 497042

Opera Account

00003490

Email

Sean.Bunn@ldh.nhs.uk

Order Number

Date Received

01/Mar/2022

Booked in By

Robert Connor

Main Company

Viamed

Type Return

For Calibration

Notes 23/Feb/2022 Zoey Teal

23/Feb/2022 Zoey Teal

Sending in a faulty VM-2500-S - needs recalibration as the CO2 accuracy is low. Serial number 60540154

01/Mar/2022 Robert Connor

Received 1 x VM-2500-S s/n 60540154, with silicone cover and 4 x AA battery

Ready For quote

agreed 28.4.22

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34749	Capnograph	4410520	60540154	N

440000 x 1 @ 1/2 hr @ £45

S/N, SRS, SRN

4420504 x 1 £1,049.75

S SRS, SRN

UPS x 1 @ £10

Reference: 116651



LUTON &
DUNSTABLE
UNIVERSITY
HOSPITAL

Work Requested: *Repair/Recalibration*

CLINICAL EXCELLENCE, QUALITY & SAFETY

Date: *23/02/21*

Hospital

Hospital	Luton & Dunstable University Hospital
Address	Lewsey Road, Luton, LU4 0DZ
Phone Number	01582 718 369
Email Address	CEWorkshop@ldh.nhs.uk

Device

Asset no	<i>21460</i>
Serial Number	<i>60540154</i>
Model	<i>UM 2500-S</i>
Description of device	<i>Capnograph</i>

Sender

Name	<i>Sean Bunn</i>
Signature	<i>Sunn</i>
Position	<i>Clinical Engineer</i>

Receiver

FAO	<i>Repairs department</i>
Company	<i>Unimed</i>
Phone	<i>NA</i>
Email	<i>NA</i>

Select Currency

Pounds| Euros| Usd

Repair SRS68087 Values in Pounds @ 1: 1.00

SRN34749	Capnograph	4410520			
	Parts Required	Reference	Quantity	Unit Cost	Total Line
		4440000	0.5	90.00	45.00
		4420504	1	1,049.75	1,049.75
				Parts Cost	1,094.75

CHECK PRICES BEFORE SENDING QUOTE

Quote Sent