

Service Repair Sheet SRS68133

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Paul James

University Hospital

Electronics Engineer

02476968386

00001288

paul.james@uhcw.nhs.uk

26/Apr/2022

Robert Connor

Viamed

Quote

Notes 14/Apr/2022 Kate Griffiths

14/Apr/2022 Kate Griffiths

Customer to send in V1000 for service/calibration. s/n GE71743941

26/Apr/2022 Robert Connor

Received 1 x V1000 s/n GE71743941, with 4 x AA battery, grey silicone cover, and blue fabric carry case.

Ready For quote

CGreen 28.4.22.

Repair Complete Signed

CGreen 6.5.22.

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35014	Foetal Simulator	1410000	GE71743941	N

1480000 x 1 @ £45
S/N, SRS, SRN

1430309 x 1 @ £0. - free
SRS, SRN,

UPS + 1 @ £10

Service Repair Sheet 68133

Contact Name

Paul James

Company/ Hospital Name

University Hospital

Position

Electronics Engineer

General Phone

02476968386

Opera Account

00001288

Email

paul.james@uhcw.nhs.uk

Order Number

200413944

Date Received

26/Apr/2022

Booked in By

Robert Connor

Repair

Ref

S/N

Equipment Type

SRN35014

1410000

GE71743941

Foetal Simulator

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1 1480000

Qty: 1 1430309

VM000 has had transducer interface cushion replaced and calibration has been checked.

Purchase Order Number : 200413944

Please quote the Purchase Order Number on all correspondence.
Payment will not be made without a valid P.O number.

Supplier :

VIAMED
15 STATION ROAD
CROSS HILLS
KEIGHLEY
WEST YORKSHIRE

BD20 7DT
Telephone : 01535 634542

Deliver To :

MEBS, Delivery Point 2
FM Building
University Hospital
Clifford Bridge Rd
Coventry
CV2 2DX

Delivery Arrangements Tel: 02476 968392

University Hospitals
Coventry and Warwickshire
NHS Trust

Invoice To :

FINANCE DEPARTMENT
University Hospitals of Coventry and Warwickshire
NHS Trust,
Clifford Bridge Road,
Coventry,
CV2 2DX

Email: accounts.payable@uhcw.nhs.uk

Order Date :

04-May-2022

Required by Date :

13-May-2022

Ordering Department :

UCP016
Biomedical Engineering

Notes to Supplier:

DELIVER TO MEBS - FM BUILDING
CONTACT PAUL JAMES 02476968386

Line No.

Quantity

Unit of
Purch

Description

Suppliers
Part No :

Contract
Reference :

Unit Price
£

Discount
£

VAT Amount
£

Line Value
£

00

1.00

SERVICE/CALIBRATION OF FOETAL HEART
SIMULATOR V1000 AS PER QUOTATION NO.
QVM136769 DATED 29.04.22

55.00

0.00

11.00

66.00

Contact in case of query :

Buyer Name : Clare Shine
Telephone No : 02476 968 417
Fax No : supplies@uhcw.nhs.uk
Email :

NHS Terms and conditions apply, a copy of which are
available on request.

VAT Excl Total : 55.00
VAT Total : 11.00
Total Order Value : 66.00