

INS

Service Repair Sheet SRS68143

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Kinga Frolyo

Company/ Hospital Name

H.S Medplan KFT (EURO)

Department

Position

Direct Phone

+36 1 361 0136

General Phone

+361 01363610272

Opera Account

00006106

Email

kinga.frolyo@medplan.hu

Order Number

Date Received

28/Apr/2022

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Warranty

Notes 26/Apr/2022 Ryan Swaine

26/Apr/2022 Ryan Swaine

3x R-22MED: 624826 Unstable in O2, 622758 low output in O2, 622759 Low output in O2. In warranty, to test and potentially replace

28/Apr/2022 Robert Connor

Received 3 x R22MED s/n 622758, 622759, 624826.

Ready For quote

€

Repair Complete Signed

Cgreen 6.5.22

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35017	Oxygen Sensor	0110022	622758	Yes
SRN35018	Oxygen Sensor	0110022	622759	Yes
SRN35019	Oxygen Sensor	0110022	624826	Yes

0140013 x 3

FOC

old S/N's, SRS, SRN's

01100022 x 3

FOC

New S/N's, SRS, SRN's

UPS x 1 FOC.