

Purchase order 12200855

Date	Part number	Part number	Serial number	Output	Zero offset voltage
	VST	JFD		mV	µV
10/05/2022	8010050	DB200381	103889	7.9	52
10/05/2022	8010050	DB200381	103890	7.9	54
10/05/2022	8010050	DB200381	103891	8.0	41
10/05/2022	8010050	DB200381	103892	7.9	50
10/05/2022	8010050	DB200381	103893	7.9	42
10/05/2022	8010050	DB200381	103894	7.9	45
10/05/2022	8010050	DB200381	103895	7.8	48
10/05/2022	8010050	DB200381	103896	8.0	62
10/05/2022	8010050	DB200381	103897	7.9	42
10/05/2022	8010050	DB200381	103898	7.9	37
10/05/2022	8010050	DB200381	103899	7.9	50
10/05/2022	8010050	DB200381	103900	7.9	40
10/05/2022	8010050	DB200381	103901	7.9	44
10/05/2022	8010050	DB200381	103902	8.0	45
10/05/2022	8010050	DB200381	103903	8.0	39
10/05/2022	8010050	DB200381	103904	8.0	47
10/05/2022	8010050	DB200381	103905	7.9	45
10/05/2022	8010050	DB200381	103906	7.8	47
10/05/2022	8010050	DB200381	103907	7.9	40
10/05/2022	8010050	DB200381	103908	8.2	43
10/05/2022	8010050	DB200381	103909	7.8	44
10/05/2022	8010050	DB200381	103910	8.0	51
10/05/2022	8010050	DB200381	103911	7.8	46
10/05/2022	8010050	DB200381	103912	7.9	36
10/05/2022	8010050	DB200381	103913	7.8	40
10/05/2022	8010050	DB200381	103914	8.0	48
10/05/2022	8010050	DB200381	103915	8.0	47
10/05/2022	8010050	DB200381	103916	8.0	53
10/05/2022	8010050	DB200381	103917	7.9	49
10/05/2022	8010050	DB200381	103918	7.9	46
10/05/2022	8010050	DB200381	103919	7.9	44
10/05/2022	8010050	DB200381	103920	7.8	44
10/05/2022	8010050	DB200381	103921	8.0	46
10/05/2022	8010050	DB200381	103922	7.9	48
10/05/2022	8010050	DB200381	103923	7.9	38
10/05/2022	8010050	DB200381	103924	8.0	40
10/05/2022	8010050	DB200381	103925	7.8	43
10/05/2022	8010050	DB200381	103926	7.9	44
10/05/2022	8010050	DB200381	103927	7.8	48
10/05/2022	8010050	DB200381	103928	7.9	44
10/05/2022	8010050	DB200381	103929	8.0	59
10/05/2022	8010050	DB200381	103930	7.8	42
10/05/2022	8010050	DB200381	103931	7.8	43
10/05/2022	8010050	DB200381	103932	7.8	54
10/05/2022	8010050	DB200381	103933	7.8	42
10/05/2022	8010050	DB200381	103934	7.9	44
10/05/2022	8010050	DB200381	103935	7.9	35
10/05/2022	8010050	DB200381	103936	8.0	35
10/05/2022	8010050	DB200381	103937	8.0	48

10/05/2022	8010050	DB200381	103938	7.9	38
10/05/2022	8010050	DB200381	103939	7.8	45
10/05/2022	8010050	DB200381	103940	8.0	58
10/05/2022	8010050	DB200381	103941	7.9	46
10/05/2022	8010050	DB200381	103942	7.9	45
10/05/2022	8010050	DB200381	103943	7.9	37
10/05/2022	8010050	DB200381	103944	7.9	38
10/05/2022	8010050	DB200381	103945	7.9	44
10/05/2022	8010050	DB200381	103946	8.0	40
10/05/2022	8010050	DB200381	103947	8.0	52
10/05/2022	8010050	DB200381	103948	7.8	41
10/05/2022	8010050	DB200381	103949	7.8	68
10/05/2022	8010050	DB200381	103950	7.8	51
10/05/2022	8010050	DB200381	103951	7.8	56
10/05/2022	8010050	DB200381	103952	7.8	38
10/05/2022	8010050	DB200381	103953	8.0	51