

Viamed Protocol for Sterilisation

1. Viamed has no facilities for sterilisation so must use a third party
2. The third party must have a validated process
3. Must produce a certificate for every batch sterilised
4. Every batch or product should be bar-coded
5. Records must be kept on every product sterilised