

Service Repair Sheet SRS68144

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Adam Murphy

Wansbeck General Hospital

Medical Electronics

Senior Medical Electronics Technician

01434655403

00000161

Adam.Murphy@northumbria-healthcare.nhs.uk

04/May/2022

Robert Connor

Viamed

Quote

Notes 27/Apr/2022 Kate Griffiths

27/Apr/2022 Kate Griffiths

Customer to send in V1000, S/N 1702630 for annual service.

04/May/2022 Robert Connor

Received 1 x V1000 s/n PR03124A11, with 4 x AA battery, hard plastic carry case and 2 x sachet of PELLJelly.

Ready For quote

CGreen 5.5.22

Repair Complete Signed

CGreen 12.5.22

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35024	Foetal Simulator	1410000	PR03124A11	N

1480000 x 1 @ £45
S/N, SRS, SRN

1430309 x 1 @ £0 free with service
SRS, SRN

UPS x 1 @ £10

Here for you...

Northumbria Healthcare
NHS Foundation Trust

NHS

Medical Electronics Department
Hexham General Hospital
Corbridge Road
Hexham
NE46 1QJ

Tel No: 0191 2934092 (Automated Answering System)

Tel No: 01434 655403

Fax No: 01434 655404

MEMO

To: Service Dept, Viamed

From: Adam Murphy, Medical Electronics, Hexham General Hospital

Date: Wednesday, 27 April 2022

Subject: Annual Service Check of V1000 Tester


Hello,

Please can you carry out the Annual Service of the V1000 Foetal Heart Simulator S/N 1702630 your returns reference SRS68144

Regards

Adam Murphy
Medical Electronics
Northumbria Healthcare NHS Foundation Trust

 Tel: 01434 655403
 Fax: 01434 655404
 Mail: Adam.Murphy@nhct.nhs.uk



**FILE
REPRINTED COPY**



Order Number :	EME2200583
----------------	------------

Requisition Number :

Date: 09/05/2022

Page: 1 of 1

Supplier		Deliver To		Invoice To			
VIAMED 15 STATION ROAD CROSS HILLS KEIGHLEY WEST YORKS		HEXHAM GENERAL HOSPITAL (ELECTRONICS) CORBRIDGE ROAD HEXHAM NORTHUMBERLAND		NORTHUMBRIA HEALTHCARE FACILITIES MGT C/O NORTHUMBRIA HOUSE UNIT 7 & 8 SILVER FOX WAY COBALT BUSINESS PARK NEWCASTLE UPON TYNE			
BD20 7DT		NE46 1QJ		NE27 0QJ			
Special Instructions :		Brief Description		CZ22 CALIBRATION OF HEART SIMULATOR AAM			
Contact/Quotation Ref :		All Enquiries to ADAM MURPHY					
Quantity	Unit	Product Description	Unit Price	VC	VAT	Total Value	Expenditure Code
		PLEASE CARRY OUT THE SERVICE OF THE HEART SIMULATOR AS PER QUOTATION QVM136841		8		55.00	552202 738200

Service Repair Sheet 68144

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Adam Murphy

Wansbeck General Hospital

Medical Electronics

Senior Medical Electronics Technician

01434655403

00000161

Adam.Murphy@northumbria-healthcare.nhs.uk

EME2200583

04/May/2022

Robert Connor

Repair	Ref	S/N	Equipment Type
SRN35024	1410000	PR03124A11	Foetal Simulator

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1 1480000

Qty: 1 1430309

V1000 has had transducer interface cushion replaced and calibration has been checked
