

Service Repair Sheet SRS67942

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Stephanie Nelle

Inspiration Medical GmbH

+49 0234 9705090

0049(0)1718361441

00006439

bestellung@inspiration-medical.com

29/Nov/2021

Robert Connor

Viamed

Quote

Notes 01/Sep/2021 Ryan Swaine

01/Sep/2021 Ryan Swaine

01/Sep/2021 Ryan Swaine

VM-2500-S, SN: A1610160012. Device beeps when turned on, but there is nothing on the display. Not under warranty, but please consider that it was only supplied to the end user in May 2021.

29/Nov/2021 Robert Connor

Received 1 x VM-2500-S s/n A1610160012

Ready For quote

C-green 22.12.2021

Repair Complete Signed

C-Green 22.12.2021

| SRN      | Equipment  | Stock Ref | Serial Number | Warranty |
|----------|------------|-----------|---------------|----------|
| SRN34371 | Capnograph | 4410520   | A1610160012   | N        |

Prices off  
Intrastats

4440000 x 0.5 @ €70€ = €35.

S/N, SRS, SRN

4400100  
~~4400100~~ x 1 @ €220.00

SRS, SRN

INS ?

← Part number  
changed.  
cg 26.5.22

UPS x 1

32x26x12cm @ 1kg