

Scheindatum: 05.11.2021
Schein-Nr.: 2021317262



INSPIRATION
Medical

Ihre Bestellung: **SRS67942**



INSPIRATION Medical GmbH Wohlfahrtstraße 115 44799 Bochum

Viamed Ltd.
15 Station Road

GBBD20 7DT Cross Hills

Lieferschein

folgende Artikel wurden wunschgemäß übergeben:

Artikelnr	Bezeichnung	Einh.	Menge	Gewicht (in kg)
IM-2500-HN	Nebenstrom Kapnograph ohne Zubehör ----- Sidestream Capnograph without accessories (REF 4410520)	1 Stk.	1,00	1,67

Our customer bought the device from us on
13.05.2021 and unfortunately already registered it
on 21.07.2021 for inspection due to a fault. It was
only in use for a few weeks.

The unit has a black screen when switched on (no
display), but when the buttons are pressed, a
beeping sound is heard.

We bought it some time ago, but our customer
bought it in May this year.

Can you please send us a cost estimate for a
repair?

Thank you very much.

If you require any further information, please feel
free to contact us Serien-Nr.: A1610160012

~~Gesamtgewicht: 1,67 kg~~

Datum: _____

Unterschrift: _____

Geschäftsführer: Christoph Böhle
Ust-IdNr.: DE276919728
Amtsgericht Bochum/HRB 13373
WEEE-Reg.-Nr. DE 94108167

Tel: +49 234 970509 0
bestellung@inspiration-medical.de
www.inspiration-medical.de

Deutsche Bank AG
BIC: DEUTDE33
IBAN: DE13430700610349000000

Service Repair Sheet 67942

Contact Name

Stephanie Nelle

Company/ Hospital Name

Inspiration Medical GmbH

Direct Phone

+49 0234 9705090

General Phone

0049(0)1718361441

Opera Account

00006439

Email

bestellung@inspiration-medical.com

Date Received

29/Nov/2021

Booked in By

Robert Connor

Repair	Ref	S/N	Equipment Type
SRN34371	4410520	A1610160012	Capnograph

Complete - Repaired
Time :0.5 Hour(s)

Parts Replaced

Qty: 0.5 4440000

Qty: 1 4400100

The Capnograph VM2500-S has been repaired. The main PCB and OLED display have been replaced.
