

Service Repair Sheet SRS68037

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Erik Sommer

Hellmut Habel Gesellschaft m.b.H.

436649657570

00006343

erik.sommer@habel-medizintechnik.at

01/Feb/2022

Robert Connor

Viamed

Warranty

Notes 06/Jan/2022 Sarah Walton

06/Jan/2022 Sarah Walton

From customer - We received the Display-Board with part number 0039030, but this one isn't working with the device. The defective Board in the device is following, after the start-up we receive an error message and the device isn't working.

S/N: T2006200029.

I have asked the customer to either just return the board or return the whole device back so that we can take a look.

01/Feb/2022 Robert Connor

Received 1 x VM-2160 s/n T2006200029

Ready For quote

CGreen 8.2.2022

Repair Complete Signed

CGreen 28.5.2022 31.5.22

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34658	Pulse Oximeter	0012165	T2006200029	Yes

Dist.

0039031 - Remove. not correct item 31.5.22
~~0039031~~ x 1 @ €100. labour + kit. - Remove
SRS, SRN
Add NOW
0039028 x 1 @ €95
SRS, SRN.
ESTD of VM2160: S/N T2006
S/N: T2006200029
0039027 x 1 @ €6.00
USP x 1 @ 0.60kg 23x15x15cm
32x24x12cm @ 0.60kg 23x15x15cm
ins?
USP x 1 @ 0.60kg 23x15x15cm

VIAMED

Return Order: SRS68037

15 Station Road, Cross Hills, Keighley
BD20 7DT WEST YORKSHIRE
ENGLAND

Ship & Invoice to:

Hellmut Habel Gesellschaft m.b.H.
Ignaz-Koeck-Strasse 20
1210 VIENNA
AUSTRIA / EUROPE
Tel.: + 43 / 1 / 292 66 42 - 0
Fax: + 43 / 1 / 290 12 45
einkauf@habel-medizintechnik.at

Enquiry no. L079251

Date: 18.01.2022

Your phone no. +44 1535 634542
Your fax no. +44 1535 635582
Your supplier no. 301782 GB287389593

DELIVERY DATE

Contact: Marion Baier
Our Customer No.: I0007025
Page: 1 / 1

Delivery terms: ex works

Shipment: UPS Standard

We herewith return for repair due to
reference number SRS68037

Part No.	Description	Qty.	Unit	Price	%	Total
99RETOURE	Viamed 1x VM-2160 Snr.: T2006200029	1,0	ea.	Lager 01		
	Failure: display defect			Lager 01		
	Return Nr. SRS68037			Lager 01		
				Lager 01		
	Value for custom: €285.-			Lager 01		
	*** Proforma Rechnung:L079251***			Lager 01		