

Service Repair Sheet SRS68118

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Chandra Suresh

Royal Free Hospital

Medical Electronics Department

Bio-Medical Engineer

0207 8302417

020 3322 1935

00003220

Chandra.Sureshkumar@nhs.net

28/Mar/2022

Robert Connor

Viamed

Quote

Notes 28/Mar/2022 Robert Connor

28/Mar/2022 Robert Connor

Received 1 x VM-2500-S s/n 81040152, with silicone cover and li-poly battery. Unit screen is smashed. Customer requests quote for screen repair, calibration, and replacement probe.

Ready For quote

CGreen 4.4.22

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34927	Capnograph	4410520	81040152	N

4440000 x 1/2 hr £90 = £45
S/N, SRS, SRN

4430020 x 1 @ £9.90 amended cg 5.4.22. Display window
SRS, SRN

4480520 x 1 @ £150
SRS, SRN

← * please Add * - cg 7.6.22. updated quote from B/Point.
UPS x 1 @ £10.
4420504 x 1 @ £1,049.75.
SRS, SRN