

Service Repair Sheet SRS68137

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Christopher Penman

Northumbria Specialist Emergency Care

01916072910

00000149

Christopher.Penman@northumbria-healthcare.nhs.uk

24/May/2022

Robert Connor

Viamed

Quote

24/05/2022 Zoey Teal

24/05/2022 Zoey Teal

24/05/2022 141000 for service and calibration. Serial number PR02569A10

24/05/2022 Robert Connor

24/05/2022 141000 s/n PR02569A10, with 4 x AA battery and hard plastic carry case.

Ready For quote

CGreen 26.5.22

Repair Complete Signed

CGNEON 10.6.22

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35060	Foetal Simulator	1410000	PR02569A10	N

1410000 x 1 @ £45

24/05/2022 SRN

1410309 x 1 @ £0 free with service

24/05/2022

1410309 x 1 @ £10

Service Repair Sheet 68137

Contact Name

Christopher Penman

Company/ Hospital Name

Northumbria Specialist Emergency Care

Direct Phone

01916072910

Opera Account

00000149

Email

Christopher.Penman@northumbria-healthcare.nhs.uk

Order Number

EME2201017

Date Received

24/May/2022

Booked in By

Robert Connor

Repair	Ref	S/N	Equipment Type
SRN35060	1410000	PR02569A10	Foetal Simulator

Complete - Repaired
Time :0 Hour(s)

Parts Replaced


Qty: 1 1480000

Qty: 1 1430309

V1000 has had transducer interface cushion replaced and calibration has been checked

PURCHASE ORDER

FILE
REPRINTED COPY



Order Number : EME2201017
Requisition Number :
Date : 30/05/2022
Page : 1 of 1

Supplier

VIAMED
15 STATION ROAD
CROSS HILLS
KEIGHLEY
WEST YORKS

BD20 7DT

Deliver To

MEDICAL ELECTRONICS DEPARTMENT (NSECH)
NORTHUMBRIA SPECIALIST EMERG CARE HOSP
MEDICAL ENGINEERING DEPARTMENT
NORTHUMBRIA WAY
CRAWLINGTON
NORTHUMBERLAND

NE23 6NZ

Invoice To

NORTHUMBRIA HEALTHCARE FACILITIES MGT
C/O NORTHUMBRIA HOUSE
UNIT 7 & 8 SILVER FOX WAY
COBALT BUSINESS PARK
NEWCASTLE UPON TYNE

NE27 0QJ

Special Instructions :
Contact/Quotation Ref :
Brief Description CZ22 V1000 SERVICE CP
All Enquiries to CHRISTOPHER PENMAN CHRISTOPHER.PENMAN@NHS.UK

Quantity	Unit	Product Description	Unit Price	VC	VAT	Total Value	Expenditure Code
		PLEASE CARRY OUT SERVICE OF V1000 AS PER QUOTATION: QVM137263 Foetal Heart Simulator V1000 - Service £45.00 V1000 Transducer interface cushion £0.00 UPS Courier Delivery - Standard £10.00 TOTAL £55.00 NOT INC VAT		8		55.00	552003 738200

Conditions of Order
1. No responsibility will be accepted for goods delivered to any point other than that specified on this order and accompanied by a Delivery Advice Note.
2. All orders will be made on this form only. The Dept will not be liable for goods or services unless ordered on this form and duly signed.
3. The Order Number above must be quoted on all advice notes, delivery notes, invoices, correspondence and acknowledgements.
4. Goods will be received only between 8am and 4pm Hrs Monday to Thursday and 8am and 3pm Friday.
5. Any alteration in Quantity or Price must be confirmed in writing by the ordering officer.
6. All goods must be accompanied by a Delivery Advice Note.

Authorised By C. G. Stans
Excluding VAT 55.00
VAT : 11.00
Including VAT 66.00