

Sterilisation.

The Apgar Timer Unit does not require sterilisation.

The instrument case can be cleaned using isopropyl alcohol.