

Service Repair Sheet SRS68141

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Ashley Davies

Thames Medical Ltd

Office Manager

01903522911

00011982

ashley@thamesmedical.com

28/Apr/2022

Robert Connor

Viamed

Quote

Notes 25/Apr/2022 Kate Griffiths

25/Apr/2022 Kate Griffiths

Customer has "IRMA (SN: 268766) that the client is reporting as making a "œwhiney noise". Unfortunately this requires an ada we do not have spare to test."

25/Apr/2022 Catherine Spence

green vets

28/Apr/2022 Robert Connor

Received 1 x IRMA s/n 268766, with lemo to minimed adapter

28/Apr/2022 Robert Connor

Correction, no adapter included, lemo connector only.

Ready For quote

Repair Complete Signed

CGreen 17.6.2022

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35016	Accessory	4420505	268766	Yes

4440010 x 1

FOC.

S/N, SRS, SRN

no fault found.

UPS x 1 FOC

Thames Medical Ltd

8 A'Becket Gardens

Worthing, West Sussex BN13 2BW

www.thamesmedical.com

email info@thamesmedical.com

tel 01903 522911

fax 01903 522922

Experts in patient monitoring and anaesthesia



# PURCHASE ORDER

order # PO-01747

date 26/04/2022

vendor address **Viamed Ltd**  
15 Station Road  
Cross Hills  
Keighley, West Yorkshire  
BD20 7DT

contact Kate  
phone 01535 634542  
fax 01535 635582

carrier  
UPS

payment terms  
Net 30

ITEM	DESCRIPTION	VENDOR PRODUCT CODE	QUANTITY	UNIT PRICE	SPECIAL TAX RATE	SUB-TOTAL
TML-V-VM2500 M1-4420505RL 2	VM2500M - IRMA C02 Analyser - Warranty Repair - Service Level 2 SN-268766	Repair of VM2500M - IRMA C02 Analyser	1	£0.00	Taxable	£0.00

remarks	SRS68141				SUB-TOTAL	£0.00
					VAT	£0.00
					TOTAL	£0.00

Please acknowledge this order and advise the delivery date by fax : +44 (0) 1903 522922 and/or by email:  
info@thamesmedical.com

Service Repair Sheet 68141

Contact Name

Company/ Hospital Name

Position

General Phone

Opera Account

Email

Date Received

Booked in By

Ashley Davies

Thames Medical Ltd

Office Manager

01903522911

00011982

ashley@thamesmedical.com

28/Apr/2022

Robert Connor

---

Repair

Ref

S/N

Equipment Type

Under Warranty

SRN35016

4420505

268766

Accessory

No Fault Found

Time :0 Hour(s)

---

The IRMA CO2 analyzer - VM2500M has been tested. The unit now works as expected.

---