

**Device Specification.**

The Device must have the ability to produce Apgar Scoring under clinical conditions.

It must have a read-out of such size as to be seen adequately from the working area of the Resuscitation Cabinet – 10mm L.C.D. display would suffice.

It must be Battery operated only (no mains outlet).

It must have sufficient back face area to accept the attachment of a Plastic Dovetail bracket for clamping to a back rail in the cabinet (dimensions are available from the Cabinet / Tom Thumb drawings).