

DVM Calibration record

DVM Calibration Index Number	
DVM Bar code	
Equipment Used	Time 1030 MicroCal
Serial number Test Equipment	6839k6
Calibration Certificate Number	1098426
Test Equipment Calibration Index Number	CE076 (Barcode 102060)
Micro Cal @	DVM Reading
0mV	
5mV	
10mV	
15mV	
20mV	
25mV	
30mV	
35mV	
40mv	
Pass / Fail	
Tested By	
Tested On	
Signed	