

Service Repair Sheet SRS68213

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Ashley Davies

Company/ Hospital Name

Thames Medical Ltd

Department

Position

Direct Phone

01903522911

General Phone

01903 522911

Opera Account

00011982

Email

ashley@thamesmedical.com

Order Number

PO-01771

Date Received

23/Jun/2022

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Quote

Notes Irma S/N 220424 ( not able to find in system under this number will re check when we receive it in)Light Flashing Red when airway adapter inserted, but reading when just blowing through without the use of the adapter

23/Jun/2022 Robert Connor

Received 1 x IRMA s/n 220424, with minimed connector. No record of this IRMA on the system.

Ready For quote

CGneen 29.6.22

Repair Complete Signed

\_\_\_\_\_

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35102	Accessory	4420505	220424	N

Dist.

4440010 x 1 FOC

S/N, SRS, SRN

unable to repair

4420505 x 1 £905

SRS, SRN

UPS x 1 @ £10

Thames Medical Ltd  
 8 A'Becket Gardens  
 Worthing, West Sussex BN13 2BW  
 www.thamesmedical.com  
 email info@thamesmedical.com  
 tel 01903 522911  
 fax 01903 522922  
 Experts in patient monitoring and anaesthesia



# PURCHASE ORDER

order # PO-01771  
 date 22/06/2022

vendor address **Viamed Ltd**  
 15 Station Road  
 Cross Hills  
 Keighley, West Yorkshire  
 BD20 7DT

contact Kate  
 phone 01535 634542  
 fax 01535 635582

carrier  
 UPS

payment terms  
 Net 30



ITEM	DESCRIPTION	VENDOR PRODUCT CODE	QUANTITY	UNIT PRICE	SPECIAL TAX RATE	SUB-TOTAL
TML-V-VM2500 M1-4420505RL 2	VM2500M - IRMA C02 Analyser - Warranty Repair - Service Level 2 SN-SN220424	Repair of VM2500M - IRMA C02 Analyser	1	£0.00	Taxable	£0.00

remarks	SRS68213	SUB-TOTAL	£0.00
	Please quote before any work takes place,	VAT	£0.00
		TOTAL	£0.00

Please acknowledge this order and advise the delivery date by fax : +44 (0) 1903 522922 and/or by email:  
 info@thamesmedical.com

Catherine Spence Required Reading 28 Issues 23 Unread 22 ISO Tasks | H/W:  
 Future Active Users 20.00 °C

▼ Search



### Select Currency

Pounds| Euros| Usd

Repair SRS68213 Values in Pounds @ £1: £1.00

SRN35102	Accessory	4420505			
	Parts Required	Reference	Quantity	Unit Cost	Total Line
		4420507	1	£0.00	£0.00
				Parts Cost	£0.00

CHECK PRICES BEFORE SENDING QUOTE

Quote Sent

Dist Price  
intra state

£905  
cg 30-6-22