

Service Repair Sheet SRS68213

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Ashley Davies

Company/ Hospital Name

Thames Medical Ltd

Department

Position

Direct Phone

01903522911

General Phone

01903 522911

Opera Account

00011982

Email

ashley@thamesmedical.com

Order Number

PO-01771

Date Received

23/Jun/2022

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Quote

Notes Irma S/N 220424 ( not able to find in system under this number will re check when we receive it in)Light Flashing Red when airway adapter inserted, but reading when just blowing through without the use of the adapter

23/Jun/2022 Robert Connor

Received 1 x IRMA s/n 220424, with minimed connector. No record of this IRMA on the system.

Ready For quote

CGneon 29.6.22

Repair Complete Signed

\_\_\_\_\_

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35102	Accessory	4420505	220424	N

Dist.

4440010 x 1 FOC

S/N, SRS, SRN

unable to repair

amended  
30.6.22 CG

4420507 x 1 £905

SRS, SRN

UPS x 1 @ £10