

OXYGEN

REPLENISHMENT CHECKLIST

AT ALL TIMES THERE MUST BE “**SIX**” CYLINDERS AVAILABLE, WITH “**ONE**” AT A TIME IN USE.

THERE SHOULD BE “**TWO**” ON THE BANK WITH “**ONE**” IN USE. WHEN THE BANK IS SWITCHED BETWEEN CYLINDERS, THE EMPTY BOTTLE **MUST** BE EXCHANGED WITH A FULL ONE.

THERE SHOULD BE “**TWO**” SPARE BOTTLES, “**ONE**” OF WHICH MAY HAVE BEEN USED.

THERE SHOULD BE “**TWO**” BOTTLES IN RESERVE. THESE SHOULD NEVER BE USED UNLESS THERE IS A WEEK OF INTENSE OXYGEN USE.

i.e. When you order, there should be “**Four**” full cylinders (including the “**Two**” in reserve)

IF YOU USE OXYGEN AND YOU CHANGE A CYLINDER, YOU ARE RESPONSIBLE FOR ENSURING THAT FRESH CYLINDERS ARE AVAILABLE, AND IF NOT, YOU **MUST RE-ORDER**

NOTE:

- FULL CYLINDERS MUST HAVE THE SEAL INTACT.
- EMPTY, OR IN-USE CYLINDERS MUST HAVE THE SEAL COMPLETELY REMOVED.