	Vandagraph Ltd.	Action Required.	Action Required	
	15,Station Road			
	Crosshills,	Date Out		
	Keighley. W.Yorks BD207DT	Invoice out		
		Comments		
Equipment				
Serial number				
Accessories	Serial Number			
Oxygen Sensor Type				
Power supply unit				
Tests C	Carried out			
Battery Check				
Soak Test				
Check Function				
Software Version				
	Calibrat	ion		
Gas	Oxygen Reading	Checks with D.P.M.		
Set 100% Oxygen	Helium Reading =	Oxygen Level	mV	
Air Reading		100.00%		
Set 100% Helium	Oxygen Reading =	75.00%		
Heliox/Trimix		50.00%		
Oxygen/Air Ratio (4.78)		25.00%		
Range 4.64 - 4.79		10.00%		
Job Carried out by		Comments		
Date Completed				
Service Time				