

 Vandagraph Ltd. 15, Station Road, Crosshills, Keighley. W.Yorks.BD207DT		Action Required	
		Customer	
		Date Out	
		Invoice Out	
		Comments	
Equipment		Calibration Equipment	
Serial number		Oxygen BOC Medical 99.95% (ML/0735/01)	
Accessories		Helium BOC Medical 99.95% (ML/0735/01)	
Oxygen Sensor Type		Heliox BOC Medical HeO ₂ 21 (ML/0735/01)	
Oxygen sensor S/N		Microcal	S/N
Power supply unit		Fluke D.V.M CE080	S/N 6911002
Power supply unit S/N		Pico ADC-200 CE020	S/N 3468
Tests Carried out			
Battery Check			
Soak Test			
Function Check			
Software version			
Calibration			
GAS		Checks with D.P.M/Microcal	
Set 100% O ₂	He =	mV	Oxygen level +/- 0.3
Air reading			100
Set 100% He	O ₂ =		75
Heliox/ Trimix	/		50
Oxygen/Air Ratio (4.78) 4.64 – 4.79			25
Job Carried out by			10