

Service Repair Sheet SRS68195

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Katharine Dixon

Sunderland Royal Hospital

0191 5656256

00004992

katharine.dixon@nhs.net

RLNP400060392

11/Jul/2022

Robert Connor

Viamed

Quote *Repair*

*Should not
be quote as
we have a
O/N...cg 18.7.22*

Notes 08/Jun/2022 Zoey Teal

08/Jun/2022 Zoey Teal

Sending in a V1000 for service and calibration. Serial number PR02010A12.

PO already received

11/Jul/2022 Robert Connor

Received 1 x V1000 s/n PR02010A12

Ready For quote

CGreen 13.7.22

Repair Complete Signed

CGreen 25.7.22

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35121	Foetal Simulator	1410000	PR02010A12	N

1480000 x 1 £145

S/N, SRS, SRN

11730309 x 1 FOC part of service

SRS, SRN

UPS x 1 £10.

RLNP BILL TO

City Hospitals Independent Commercial Enterprises Ltd C/O
Financial Services South Tyneside District Hospital, Harton Lane
NE34 0PL
South Shields
GLN: 5055235808131 (GLN, Endpoint ID)

Contact details
Katharine Dixon
Email: katharine.dixon@nhs.uk

Order nbr

RLNP400060392

Date: 2022-06-08

Validity:

Seller
Viamed Ltd
BD207DT Cross Hills,
Keighley
Country: United
Kingdom
GLN: 5051826000007
(GLN, Endpoint ID)
GLN:
5051826000007(0088)

Delivery

90p71a - Electronics Department (Ebme) A-Floor,
Sunderland Royal Hospital Kayll Road
SR4 7TP Sunderland
Country: United Kingdom
Delivery party: 1 90P71A - ELECTRONICS DEPT
(EBME) (CHOICE LTD)
Delivery place: 1 90P71A - ELECTRONICS DEPT
(EBME) (CHOICE LTD)

Recipient

**City Hospitals Independent
Commercial Enterprises Ltd**
Sunderland Royal Hospital Kayll
Road
Address:
SR4 7TP Sunderland
VAT Reg No: GB180373709 (9932
Endpoint ID)

Line	Product	Description	Quantity	Unit price	Total
#1	SERVICE & CALIBRATION	SERVICE & CALIBRATION SERVICE & CALIBRATION OF A VIAMED V1000 (SERIAL NO: PRO2010A12)	1.0 EA	45.00	45.00 GBP
#2	POSTAGE	POSTAGE POSTAGE & PACKAGING	1.0 EA	10.00	10.00 GBP

Service Repair Sheet 68195

Contact Name

Katharine Dixon

Company/ Hospital Name

Sunderland Royal Hospital

General Phone

0191 5656256

Opera Account

00004992

Email

katharine.dixon@nhs.net

Order Number

RLNP400060392

Date Received

11/Jul/2022

Booked in By

Robert Connor

Repair

Ref

S/N

Equipment Type

SRN05121

1410000

PR02010A12

Foetal Simulator

Complete - Repaired

Time :0 Hour(s)

Parts Replaced

Qty: 1 1480000

Qty: 1 1430309

1410000 has had transducer interface cushion replaced and calibration has been checked