

Service Repair Sheet SRS68208

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Miroslav Trebicky

ATAL s.r.o.

00420 381251791

420381251791

00006059

trebicky@atal.cz

20/Jul/2022

Robert Connor

Viamed

Warranty

Notes 20/Jun/2022 Zoey Teal

20/Jun/2022 Zoey Teal

Returning 2 R-22A sensors.

Serial number 663097 - does not fall below 0.5.

Serial number 774334 - without voltage

20/Jul/2022 Robert Connor

Received 3 x R22A s/n 663097, 774334, 774516.

Ready For quote

CGNEON 25.7.2022

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35136	Oxygen Sensor	0110122	774334	Yes
SRN35135	Oxygen Sensor	0110122	663097	Yes
SRN35137	Oxygen Sensor	0110122	774516	Yes

0140013 x 3
old
S/N's, SRS, SEN's

FOC

0110122 x 3
New S/N's, SRS, SEN

FOC.

UPS x 1 FOC.

23.15.2022 0.40kg.



ATAL s.r.o.

Lesní 47, 390 01 Tábor - Horky, Czech Republic

Tel.: + 420 381 410 121

DELIVERY NOTE

Nr. Q06/2022

Datum: 11.7.22

RMA: SRS68208

addressee

Viamed Limited

15 Station Road

BD20 7DT Cross Hills

Keighley

West Yorkshire UK

Zoey Teal +44 (0)1535 634542

Reference	Description	S/N	
1. R-22A	oxygen sensor	663097	the value O2 does not fall below 0,5
2. R-22A	oxygen sensor	774334	without voltage
3. R-22A	oxygen sensor	774516	low voltage

OAS 00915 - 17 -

GOODS TO REPAIR OF WARRANTY

INCOTERMS: DAP UPS
Parcel quantity: 1
Weight: 0,25 kg
25x19x6 cm

Issue: Miroslav Třebický

trebicky@actia.cz

Confirmation of consumer:

Take-over date:

Confirmed Delivery note send back to our address (Please at e-mail too)!

Service Repair Sheet 68208

Contact Name	Miroslav Trebicky
Company/ Hospital Name	ATAL s.r.o.
Direct Phone	00420 381251791
General Phone	420381251791
Opera Account	00006059
Email	trebicky@atal.cz
Date Received	20/Jul/2022
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN35135	0110122	663097	Oxygen Sensor	

Time :0 Hour(s)

Parts Replaced

Qty: 1 0110122

Replaced under warranty with S/N 829803

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN35136	0110122	774334	Oxygen Sensor	

Time :0 Hour(s)

Parts Replaced

Qty: 1 0110122

Replaced under warranty with S/N 829802

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN35137	0110122	774516	Oxygen Sensor	

Time :0 Hour(s)

Parts Replaced

Qty: 1 0110122

Replaced under warranty with S/N 829804
