

Service Repair Sheet SRS68208

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Miroslav Trebicky

ATAL s.r.o.

00420 381251791

420381251791

00006059

trebicky@atal.cz

20/Jul/2022

Robert Connor

Viamed

Warranty

Notes 20/Jun/2022 Zoey Teal

20/Jun/2022 Zoey Teal

Returning 2 R-22A sensors.

Serial number 663097 - does not fall below 0.5.

Serial number 774334 - without voltage

20/Jul/2022 Robert Connor

Received 3 x R22A s/n 663097, 774334, 774516.

Ready For quote

CGNEON 25.7.2022

Repair Complete Signed

CGNEON 26.7.22

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35136	Oxygen Sensor	0110122	774334	Yes
SRN35135	Oxygen Sensor	0110122	663097	Yes
SRN35137	Oxygen Sensor	0110122	774516	Yes

0140013 x 3
old
S/N's, SRS, SRN's

FOC

0110122 x 3
New S/N's, SRS, SRN

FOC.

UPS x 1 FOC.

23x15x15cm 0.40kg.