

Service Repair Sheet SRS68222

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Joshua Mantle

Company/ Hospital Name

Chesterfield Royal Hospital

Department

Clinical Engineering

Position

Clinical Engineer

Direct Phone

01246 512058

General Phone

01246 277 271

Opera Account

00001170

Email

joshua.mantle@nhs.net

Order Number

80061093

Date Received

26/Jul/2022

Booked in By

Robert Connor

Main Company

Viamed

Type Return

For Servicing

Notes 08/Jul/2022 Aqib Majeed

08/Jul/2022 Sophie Lines

Customer has sent PO for V1000 calibration. PO attached - Goods not yet received

08/Jul/2022 Sophie Lines

Serial number: PRO408A12

26/Jul/2022 Robert Connor

Received 1 x V1000 s/n PR0408A12, with 4 x AA battery and blue fabric case.

Ready For quote

Repair Complete Signed

CC Neon 27.7.22

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35153	Foetal Simulator	1410000	PR0408A12	N

1480000 x 1 £45
S/N, SRS, SRN

1430309 x 1 £0 FOC
SRS, SRN

GPS x 1 @ £10

Service Repair Sheet 68222

Contact Name Joshua Mantle
Company/ Hospital Name Chesterfield Royal Hospital
Department Clinical Engineering
Position Clinical Engineer
Direct Phone 01246 512058
General Phone 01246 277 271
Opera Account 00001170
Email joshua.mantle@nhs.net
Order Number 80061093
Date Received 26/Jul/2022
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	Complete - Repaired
SRN35153	1410000	PR0408A12	Foetal Simulator	Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1480000	
		Qty: 1	1430309	

V1000 has had transducer interface cushion replaced and calibration has been checked



Derbyshire Support
and Facilities Services

Purchase Order

VIAMED LTD
15 STATION ROAD
CROSSHILLS
KEIGHLEY
WEST YORKSHIRE
BD20 7DT

Order number 80061093
Order date 08/07/2022
Delivery attention
Payment terms 30 Days
Your ref Calibration of V1000
Foetal Heart Simulator

Deliver to:
Chesterfield Royal Hospital
Receipts & Distribution
Top Road, Calow, Chesterfield, S44 5BL

Invoice to:
DSFS Limited
c/o Chesterfield Royal Hospital
Financial Services Department, Top Road, Calow,
Chesterfield, Derbyshire, S44 5BL

Tel No. 01246 277271
Fax No. 01246 512610
Email: CRHFT.Procurement@nhs.net

Email: crhft.dsfs.pdfinvoices@nhs.net

Our opening hours are 9am - 4pm
All correspondence must contain our order number

Page 1 of

Article	Description	Qty	Unit	Unit Price	Amount in GBP	Delivery Date
1115961	Clinical Engineering Spare Parts [Please see details] Calibration of V1000 Foetal Heart Simulator - SN: PRO408A12Viamed Part Number: 1480000. Price: £45.00	1.00	EA	45.00	45.00	14/07/2022
1115961	Clinical Engineering Spare Parts [Please see details] Carriage Charge - £10.00	1.00	EA	10.00	10.00	14/07/2022

Subject to NHS Model terms and conditions for purchase of goods and services, unless otherwise stated.

Total Net 55.00
Total VAT 11.00
Total in GBP 66.00

Derbyshire Support and Facilities Services Limited is a wholly-owned subsidiary of Chesterfield Royal Hospital NHS Foundation Trust. Companies House Registration: 11494436. Registered Office : Finance Department, Chesterfield Royal Hospital, Calow, Chesterfield, Derbyshire S44 5BL