

Device Safety Classification

SpO2 Probes have been tested to DIN 60601-1 /03.96/

Measurements for Device Safety – Report No 498ENV12

20th January 1998 - EnviteC – Wismar – Germany.

The tests showed no relevant function impairment

The test results of the above can be found in Section “Z – Sub-section 20”