

Service Repair Sheet SRS68241

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

David Bell

Northumbria Specialist Emergency

Medical Engineering Department

00000149

DAVID.BELL2@NHCT.NHS.UK

EME2202219

15/Aug/2022

Robert Connor

Viamed

~~Quote~~ Service

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☐

Cleaned by Viamed,
if no declaration certificate
from customer ☒

Signed: RC
Date: 15/8/22

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____
Date: _____

Notes 01/Aug/2022 Aqib Majeed
01/Aug/2022 Aqib Majeed
customer sending v1000 for service.
No sn number given.
see po attached

15/Aug/2022 Robert Connor
Received 1 x V1000 s/n PR03124A10, with 4 x AA battery. For annual service/calibration.

Ready For quote

Repair Complete Signed

CCNEON 15.8.22

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35201	Foetal Simulator	1410000	PR03124A10	N

1480000 x 1 @ £45

S/N, SRS, SRN

1430309 x 1 @ £0 - free with service

SRS, SRN

UPS x 1 @ £10

Service Repair Sheet 68241

Contact Name

Company/ Hospital Name

Department

Opera Account

Email

Order Number

Date Received

Booked in By

David Bell

Northumbria Specialist Emergency Care

Medical Engineering Department

00000149

DAVID.BELL2@NHCT.NHS.UK

EME2202219

15/Aug/2022

Robert Connor

Repair	Ref	S/N	Equipment Type
SRN35201	1410000	PR03124A10	Foetal Simulator

Complete - Repaired
Time :0 Hour(s)


Parts Replaced

Qty: 1 1480000

Qty: 1 1430309

V1000 has had transducer interface cushion replaced and calibration has been checked

PURCHASE ORDER



**FILE
REPRINTED COPY**

Order Number: EME2202219
Requisition Number:
Date: 28/07/2022
Page: 1 of 1

Supplier	Deliver To	Invoice To
VIAMED 15 STATION ROAD CROSS HILLS KEIGHLEY WEST YORKS	MEDICAL ELECTRONICS DEPARTMENT (NSECH) NORTHUMBRIA SPECIALIST EMERG CARE HOSP MEDICAL ENGINEERING DEPARTMENT NORTHUMBRIA WAY CRAWLINGTON NORTHUMBERLAND	NORTHUMBRIA HEALTHCARE FACILITIES MGT C/O NORTHUMBRIA HOUSE UNIT 7 & 8 SILVER FOX WAY COBALT BUSINESS PARK NEWCASTLE UPON TYNE
BD20 7DT	NE23 6NZ	NE27 0QJ

Special Instructions: Brief Description CZ22 V1000 SERVICE DB
Contact/Quotation Ref: All Enquiries to DAVE BELL DAVID.BELL2@NHCT.NHS.UK

Quantity	Unit	Product Description	Unit Price	VC	VAT	Total Value	Expenditure Code
		PLEASE CARRY OUT SERVICE OF V1000		8		55.00	552052 738200
		Foetal Heart Simulator V1000 - Service £45.00 V1000 Transducer interface cushion £0.00 UPS Courier Delivery - Standard £10.00 TOTAL £55.00 NOT INC VAT					

Conditions of Order
1. No responsibility will be accepted for goods delivered to any point other than that specified on this order and accompanied by a Delivery Advice Note.
2. All orders will be made on this form only. The Dept will not be liable for goods or services unless ordered on this form and duly signed.
3. The Order Number above must be quoted on all advice notes, delivery notes, invoices, correspondence and acknowledgements.
4. Goods will be received only between 8am and 4pm Hrs Monday to Thursday and 8am and 3pm Friday.
5. Any alteration in Quantity or Price must be confirmed in writing by the ordering officer.
6. All goods must be accompanied by a Delivery Advice Note.

Authorised By: *C. Glass*
CHRIS GLASS