

EU - Conformity Statement

We declare with sole responsibility that following product comply with the basic requirements according to the Council Directive 93/42/EEC (Annex I) concerning medical devices.

Product: **Oxygen Sensor**

Typ:

OOM101	OOM104	OOM110	OOM201
OOM102	OOM105	OOM111	OOM202
OOM102-1	OOM106	OOM112	OOM202-1
OOM103	OOM107		OOM202-2
OOM103-1	OOM107-2		OOM202-2S
OOM103-1M			OOM204

Classification:
(RL 93/42/EEC, Annex IX)

Class IIa

CE marking:



Notified Body:

TÜV SÜD Product Service GmbH, Munic, Germany

Conformity Assessment
Process:

Annex II, section 3 of the Directive 93/42/EEC

Particular product
standard applied:

EN ISO 21647:2009

Issued by:

ENVITEC-WISMAR GMBH
Alter Holzhafen 18
D-23966 Wismar
Germany

Valid from:

2010-09-09

Place, Date:

Wismar, 2010-09-09

Authorized Signature:

Marcus Lindenlaub
Managing Director